

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010920\
 Data File : BE101017.D
 Acq On : 9 Jan 2020 18:07
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 10 01:03:07 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA E\Methods\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	-0.01
2	1,4-Dioxane	1.020	0.984	3.5	97	0.00
3	n-Nitrosodimethylamine	0.556	0.454	18.3	88	0.00
4 S	2-Fluorophenol	0.900	0.805	10.6	84	0.00
5 S	Phenol-d6	1.139	0.939	17.6	86	0.00
6	bis(2-Chloroethyl)ether	1.357	1.118	17.6	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.286	0.212	25.9#	79	0.00
9	Naphthalene	4.441	4.591	-3.4	96	0.00
10	Hexachlorobutadiene	0.188	0.197	-4.8	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.763	0.777	-1.8	96	0.00
12	2-Methylnaphthalene	0.672	0.657	2.2	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.113	0.078	31.0#	76	0.00
15 S	2-Fluorobiphenyl	1.519	1.581	-4.1	98	0.00
16	Acenaphthylene	1.629	1.370	15.9	89	0.00
17	Acenaphthene	1.166	1.140	2.2	96	0.00
18	Fluorene	1.474	1.430	3.0	94	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.195	0.182	6.7	92	0.00
21	Hexachlorobenzene	0.214	0.225	-5.1	98	0.00
22	Pentachlorophenol	0.053	0.029	45.3#	67	0.00
23	Phenanthrene	1.092	1.068	2.2	93	0.00
24	Anthracene	1.031	0.952	7.7	91	0.00
25 SURR	Fluoranthene-d10	5.555	5.100	8.2	89	0.00
26	Fluoranthene	1.368	1.232	9.9	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	-0.01
28	Pvrene	1.489	1.490	-0.1	88	0.00
29 S	Terphenyl-d14	0.982	0.968	1.4	88	0.00
30	Benzo(a)anthracene	1.141	0.922	19.2	76	0.00
31	Chrysene	1.696	1.931	-13.9	100	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	1.591	16.4	87	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.372	-2.4	95	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	93	0.00
35	Benzo(b)fluoranthene	1.122	1.006	10.3	89	0.00
36	Benzo(k)fluoranthene	1.626	1.460	10.2	83	0.00
37 C	Benzo(a)pyrene	1.085	1.012	6.7	89	0.00
38	Dibenzo(a,h)anthracene	0.998	0.885	11.3	89	-0.01
39	Benzo(a,h,i)perylene	1.227	1.232	-0.4	90	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010920\
Data File : BE101017.D
Acq On : 9 Jan 2020 18:07
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 10 01:03:07 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA E\Methods\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		