

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011120\
 Data File : BE101033.D
 Acq On : 10 Jan 2020 21:20
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 11 01:48:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	1.020	0.984	3.5	101	0.00
3	n-Nitrosodimethylamine	0.556	0.430	22.7	87	0.00
4 S	2-Fluorophenol	0.900	0.792	12.0	86	0.00
5 S	Phenol-d6	1.139	0.930	18.3	89	0.00
6	bis(2-Chloroethyl)ether	1.357	1.039	23.4	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.286	0.163	43.0#	69	0.01
9	Naphthalene	4.441	4.342	2.2	103	0.00
10	Hexachlorobutadiene	0.188	0.187	0.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.763	0.749	1.8	105	0.00
12	2-Methylnaphthalene	0.672	0.632	6.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.113	0.068	39.8#	73	0.00
15 S	2-Fluorobiphenyl	1.519	1.513	0.4	104	0.00
16	Acenaphthylene	1.629	1.268	22.2	91	0.00
17	Acenaphthene	1.166	1.143	2.0	106	0.00
18	Fluorene	1.474	1.421	3.6	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.195	0.178	8.7	98	0.00
21	Hexachlorobenzene	0.214	0.212	0.9	100	0.00
22	Pentachlorophenol	0.053	0.026	50.9#	64	0.00
23	Phenanthrene	1.092	1.031	5.6	98	0.00
24	Anthracene	1.031	0.932	9.6	97	0.00
25 SURR	Fluoranthene-d10	5.555	4.992	10.1	95	0.00
26	Fluoranthene	1.368	1.219	10.9	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	-0.01
28	Pvrene	1.489	1.258	15.5	94	0.00
29 S	Terphenyl-d14	0.982	0.814	17.1	93	0.00
30	Benzo(a)anthracene	1.141	0.984	13.8	102	0.00
31	Chrysene	1.696	1.888	-11.3	123	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.361	-24.1	162#	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.547	-15.4	135	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	127	-0.01
35	Benzo(b)fluoranthene	1.122	0.909	19.0	110	-0.01
36	Benzo(k)fluoranthene	1.626	1.500	7.7	117	0.00
37 C	Benzo(a)pyrene	1.085	1.080	0.5	130	0.00
38	Dibenzo(a,h)anthracene	0.998	0.887	11.1	122	-0.02
39	Benzo(a,h,i)perylene	1.227	1.174	4.3	118	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011120\
Data File : BE101033.D
Acq On : 10 Jan 2020 21:20
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 11 01:48:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			