

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011420\
 Data File : BE101047.D
 Acq On : 14 Jan 2020 9:37
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 15 10:03:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.01
2	1,4-Dioxane	1.020	1.001	1.9	104	0.00
3	n-Nitrosodimethylamine	0.556	0.410	26.3#	84	0.01
4 S	2-Fluorophenol	0.900	0.783	13.0	86	0.00
5 S	Phenol-d6	1.139	1.031	9.5	100	0.00
6	bis(2-Chloroethyl)ether	1.357	1.373	-1.2	109	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.286	0.232	18.9	96	0.01
9	Naphthalene	4.441	4.553	-2.5	106	0.00
10	Hexachlorobutadiene	0.188	0.202	-7.4	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.763	0.787	-3.1	109	0.00
12	2-Methylnaphthalene	0.672	0.666	0.9	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.113	0.063	44.2#	70	0.00
15 S	2-Fluorobiphenyl	1.519	1.562	-2.8	109	0.00
16	Acenaphthylene	1.629	1.284	21.2	94	0.00
17	Acenaphthene	1.166	1.149	1.5	109	-0.01
18	Fluorene	1.474	1.416	3.9	106	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.195	0.180	7.7	101	0.00
21	Hexachlorobenzene	0.214	0.227	-6.1	109	0.00
22	Pentachlorophenol	0.053	0.025	52.8#	64	0.00
23	Phenanthrene	1.092	1.073	1.7	103	0.00
24	Anthracene	1.031	0.939	8.9	100	0.00
25 SURR	Fluoranthene-d10	5.555	4.972	10.5	96	0.00
26	Fluoranthene	1.368	1.227	10.3	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	-0.01
28	Pvrene	1.489	1.625	-9.1	94	0.00
29 S	Terphenyl-d14	0.982	1.051	-7.0	94	0.00
30	Benzo(a)anthracene	1.141	0.876	23.2	71	0.00
31	Chrysene	1.696	1.903	-12.2	97	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	1.299	31.7#	70	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.370	-2.2	93	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.122	0.920	18.0	86	0.00
36	Benzo(k)fluoranthene	1.626	1.548	4.8	93	0.00
37 C	Benzo(a)pyrene	1.085	0.987	9.0	92	0.00
38	Dibenzo(a,h)anthracene	0.998	0.882	11.6	94	-0.01
39	Benzo(a,h,i)perylene	1.227	1.169	4.7	91	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011420\
Data File : BE101047.D
Acq On : 14 Jan 2020 9:37
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 15 10:03:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			