

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011519\
 Data File : BE098632.D
 Acq On : 15 Jan 2019 17:48
 Operator : SJ/JU
 Sample : K1029-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 50

Instrument :
 BNA_E
 ClientSampleId :
 B12_123118

Manual Integrations
 APPROVED

Sohil
 1/16/2019 4:05:37 PM

Quant Time: Jan 16 10:00:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	60631	20.00	ng	0.00
7) Naphthalene-d8	10.60	136	237802	20.00	ng	0.00
13) Acenaphthene-d10	14.47	164	151004	20.00	ng	0.00
19) Phenanthrene-d10	17.23	188	358492	20.00	ng	0.00
27) Chrysene-d12	21.41	240	425274	20.00	ng	0.00
34) Perylene-d12	23.92	264	436978	20.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.38	112	186934	57.80	ng	-0.01
5) Phenol-d6	6.97	99	168835	36.66	ng	-0.02
8) Nitrobenzene-d5	8.97	82	460106	121.97	ng	0.00
11) 2-Methylnaphthalene-d10	12.43	152	67	0.01	ng	0.22
14) 2,4,6-Tribromophenol	15.97	330	216959	165.11	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	881400	71.26	ng	0.00
25) Fluoranthene-d10	19.25	212	27	0.00	ng	-0.01
29) Terphenyl-d14	19.86	244	1531973	81.40	ng	0.00
Target Compounds						
30) Benzo(a)anthracene	21.40	228	1532m	0.067	ng	Qvalue
32) Bis(2-ethylhexyl)phthalate	21.32	149	21154	0.436	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011519\
Data File : BE098632.D
Acq On : 15 Jan 2019 17:48
Operator : SJ/JU
Sample : K1029-01
Misc :
ALS Vial : 13 Sample Multiplier: 50

Instrument :
BNA_E
ClientSampleId :
B12_123118

Manual Integrations
APPROVED

Sohil
1/16/2019 4:05:37 PM

Quant Time: Jan 16 10:00:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

