

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011519\
 Data File : BE098634.D
 Acq On : 15 Jan 2019 19:00
 Operator : SJ/JU
 Sample : K1029-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 50

Instrument :
 BNA_E
 ClientSampleId :
 MW12 123118

Manual Integrations
 APPROVED

Sohil
 1/16/2019 4:05:40 PM

Quant Time: Jan 16 10:09:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	59165	20.00	ng	0.00
7) Naphthalene-d8	10.60	136	235974	20.00	ng	0.00
13) Acenaphthene-d10	14.47	164	149687	20.00	ng	0.00
19) Phenanthrene-d10	17.23	188	361648	20.00	ng	0.00
27) Chrysene-d12	21.42	240	406997	20.00	ng	0.00
34) Perylene-d12	23.92	264	414350	20.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	201740	63.92	ng	0.00
5) Phenol-d6	6.98	99	180816	40.23	ng	-0.02
8) Nitrobenzene-d5	8.96	82	585042	156.29	ng	0.00
11) 2-Methylnaphthalene-d10	12.21	152	40	0.01	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	285187	218.94	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	1160115	94.62	ng	0.01
25) Fluoranthene-d10	19.26	212	21	0.00	ng	0.00
29) Terphenyl-d14	19.86	244	1993610	110.69	ng	0.00
Target Compounds						
22) Pentachlorophenol	16.89	266	50	10.215	ng	# 80
23) Phenanthrene	17.27	178	693	0.040	ng	# 83
30) Benzo(a)anthracene	21.40	228	1418m	0.065	ng	
32) Bis(2-ethylhexyl)phthalate	21.32	149	12263	0.264	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011519\
Data File : BE098634.D
Acq On : 15 Jan 2019 19:00
Operator : SJ/JU
Sample : K1029-05
Misc :
ALS Vial : 15 Sample Multiplier: 50

Instrument :
BNA_E
ClientSampleId :
MW12 123118

Manual Integrations
APPROVED

Sohil
1/16/2019 4:05:40 PM

Quant Time: Jan 16 10:09:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

