

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
 Data File : BE098649.D
 Acq On : 17 Jan 2019 12:58
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jan 17 13:33:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.02
2	1,4-Dioxane	0.400	0.397	0.8	90	0.00
3	n-Nitrosodimethylamine	0.400	0.458	-14.5	107	0.00
4 S	2-Fluorophenol	0.400	0.423	-5.7	91	0.01
5 S	Phenol-d6	0.400	0.401	-0.3	84	0.01
6	bis(2-Chloroethyl)ether	0.400	0.422	-5.5	96	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	84	0.02
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	93	0.02
9	Naphthalene	0.400	0.385	3.8	86	0.01
10	Hexachlorobutadiene	0.400	0.391	2.3	87	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	89	0.01
12	2-Methylnaphthalene	0.400	0.377	5.8	86	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.01
14 S	2,4,6-Tribromophenol	0.400	0.410	-2.5	94	0.01
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	91	0.03
16	Acenaphthylene	0.400	0.375	6.3	91	0.01
17	Acenaphthene	0.400	0.367	8.3	87	0.01
18	Fluorene	0.400	0.370	7.5	87	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4-Bromophenyl-phenylether	0.400	0.358	10.5	86	0.01
21	Hexachlorobenzene	0.400	0.376	6.0	88	0.01
22	Pentachlorophenol	0.400	0.471	-17.7	107	0.01
23	Phenanthrene	0.400	0.382	4.5	91	0.00
24	Anthracene	0.400	0.376	6.0	93	0.00
25 SURR	Fluoranthene-d10	0.400	0.376	6.0	90	0.00
26	Fluoranthene	0.400	0.369	7.8	89	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
28	Pvrene	0.400	0.375	6.3	88	0.00
29 S	Terphenyl-d14	0.400	0.379	5.3	92	0.00
30	Benzo(a)anthracene	0.400	0.368	8.0	87	0.01
31	Chrysene	0.400	0.372	7.0	88	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.346	13.5	87	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.385	3.8	89	0.00
34 I	Perylene-d12	0.400	0.400	0.0	90	0.00
35	Benzo(b)fluoranthene	0.400	0.372	7.0	90	0.00
36	Benzo(k)fluoranthene	0.400	0.382	4.5	89	0.00
37 C	Benzo(a)pyrene	0.400	0.371	7.3	89	0.00
38	Dibenzo(a,h)anthracene	0.400	0.397	0.8	95	0.00
39	Benzo(a,h,i)perylene	0.400	0.379	5.3	89	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
Data File : BE098649.D
Acq On : 17 Jan 2019 12:58
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 17 13:33:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				