

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
 Data File : BE098651.D
 Acq On : 17 Jan 2019 16:24
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 18 02:00:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.789	0.746	5.4	88	0.00
3	n-Nitrosodimethylamine	0.579	0.561	3.1	82	0.01
4 S	2-Fluorophenol	1.067	0.994	6.8	77	0.00
5 S	Phenol-d6	1.519	1.532	-0.9	81	0.00
6	bis(2-Chloroethyl)ether	1.149	1.110	3.4	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
8 S	Nitrobenzene-d5	0.317	0.310	2.2	83	0.01
9	Naphthalene	3.984	3.654	8.3	80	0.00
10	Hexachlorobutadiene	0.285	0.284	0.4	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.620	5.8	84	0.00
12	2-Methylnaphthalene	0.660	0.610	7.6	82	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.174	0.166	4.6	84	0.00
15 S	2-Fluorobiphenyl	1.638	1.459	10.9	85	0.00
16	Acenaphthylene	1.732	1.573	9.2	84	0.00
17	Acenaphthene	1.087	0.985	9.4	82	0.00
18	Fluorene	1.414	1.313	7.1	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4-Bromophenyl-phenylether	0.212	0.193	9.0	85	0.00
21	Hexachlorobenzene	0.256	0.231	9.8	81	0.00
22	Pentachlorophenol	0.075	0.070	6.7	102	0.00
23	Phenanthrene	0.948	0.846	10.8	82	0.00
24	Anthracene	0.812	0.688	15.3	81	0.00
25 SURR	Fluoranthene-d10	4.955	4.596	7.2	86	0.00
26	Fluoranthene	1.220	1.109	9.1	85	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	-0.01
28	Pvrene	1.101	0.979	11.1	85	0.00
29 S	Terphenyl-d14	0.885	0.773	12.7	87	0.00
30	Benzo(a)anthracene	1.075	0.973	9.5	87	0.00
31	Chrysene	1.115	0.953	14.5	83	0.00
32	Bis(2-ethylhexyl)phthalate	2.281	1.956	14.2	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.150	1.047	9.0	86	0.00
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	1.091	0.952	12.7	86	0.00
36	Benzo(k)fluoranthene	1.193	1.077	9.7	86	0.00
37 C	Benzo(a)pyrene	1.072	0.952	11.2	86	0.00
38	Dibenzo(a,h)anthracene	0.949	0.881	7.2	91	0.00
39	Benzo(a,h,i)perylene	1.040	0.935	10.1	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
Data File : BE098651.D
Acq On : 17 Jan 2019 16:24
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 18 02:00:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			