

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
 Data File : BE098661.D
 Acq On : 17 Jan 2019 23:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 18 01:58:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.789	0.620	21.4	80	0.00
3	n-Nitrosodimethylamine	0.579	0.557	3.8	89	0.00
4 S	2-Fluorophenol	1.067	1.040	2.5	89	0.00
5 S	Phenol-d6	1.519	1.529	-0.7	89	0.00
6	bis(2-Chloroethyl)ether	1.149	1.054	8.3	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.317	0.367	-15.8	103	0.00
9	Naphthalene	3.984	3.606	9.5	83	0.00
10	Hexachlorobutadiene	0.285	0.277	2.8	88	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.617	6.2	87	0.00
12	2-Methylnaphthalene	0.660	0.598	9.4	85	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.174	0.182	-4.6	96	0.00
15 S	2-Fluorobiphenyl	1.638	1.563	4.6	94	0.00
16	Acenaphthylene	1.732	1.567	9.5	87	0.00
17	Acenaphthene	1.087	0.973	10.5	84	0.00
18	Fluorene	1.414	1.304	7.8	86	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	-0.01
20	4-Bromophenyl-phenylether	0.212	0.194	8.5	90	0.00
21	Hexachlorobenzene	0.256	0.231	9.8	86	0.00
22	Pentachlorophenol	0.075	0.064	14.7	100	0.00
23	Phenanthrene	0.948	0.832	12.2	86	0.00
24	Anthracene	0.812	0.697	14.2	87	0.00
25 SURR	Fluoranthene-d10	4.955	4.302	13.2	86	0.00
26	Fluoranthene	1.220	1.055	13.5	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.101	0.969	12.0	84	0.00
29 S	Terphenyl-d14	0.885	0.843	4.7	95	0.00
30	Benzo(a)anthracene	1.075	0.951	11.5	86	0.00
31	Chrysene	1.115	0.987	11.5	86	0.00
32	Bis(2-ethylhexyl)phthalate	2.281	1.901	16.7	86	0.00
33	Indeno(1,2,3-cd)pyrene	1.150	1.069	7.0	89	0.00
34 I	Perylene-d12	1.000	1.000	0.0	93	0.00
35	Benzo(b)fluoranthene	1.091	0.987	9.5	90	0.00
36	Benzo(k)fluoranthene	1.193	1.082	9.3	87	0.00
37 C	Benzo(a)pyrene	1.072	0.992	7.5	91	0.00
38	Dibenzo(a,h)anthracene	0.949	0.923	2.7	96	-0.01
39	Benzo(a,h,i)perylene	1.040	0.955	8.2	88	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011719\
Data File : BE098661.D
Acq On : 17 Jan 2019 23:55
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 18 01:58:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			