

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011819\
 Data File : BE098663.D
 Acq On : 18 Jan 2019 8:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 18 11:37:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 15 14:30:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.789	0.734	7.0	87	0.00
3	n-Nitrosodimethylamine	0.579	0.494	14.7	72	0.00
4 S	2-Fluorophenol	1.067	1.057	0.9	82	0.00
5 S	Phenol-d6	1.519	1.504	1.0	80	0.00
6	bis(2-Chloroethyl)ether	1.149	1.055	8.2	80	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
8 S	Nitrobenzene-d5	0.317	0.341	-7.6	93	0.01
9	Naphthalene	3.984	3.631	8.9	81	0.00
10	Hexachlorobutadiene	0.285	0.283	0.7	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.602	8.5	82	0.00
12	2-Methylnaphthalene	0.660	0.596	9.7	81	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.174	0.168	3.4	87	0.00
15 S	2-Fluorobiphenyl	1.638	1.383	15.6	83	0.01
16	Acenaphthylene	1.732	1.526	11.9	84	0.00
17	Acenaphthene	1.087	0.957	12.0	82	0.00
18	Fluorene	1.414	1.285	9.1	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.212	0.184	13.2	83	0.00
21	Hexachlorobenzene	0.256	0.229	10.5	83	0.00
22	Pentachlorophenol	0.075	0.076	-1.3	114	0.00
23	Phenanthrene	0.948	0.844	11.0	84	0.00
24	Anthracene	0.812	0.674	17.0	81	0.00
25 SURR	Fluoranthene-d10	4.955	4.472	9.7	86	0.00
26	Fluoranthene	1.220	1.082	11.3	85	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pvrene	1.101	0.975	11.4	83	0.00
29 S	Terphenyl-d14	0.885	0.784	11.4	87	0.00
30	Benzo(a)anthracene	1.075	0.961	10.6	86	0.00
31	Chrysene	1.115	0.993	10.9	85	0.00
32	Bis(2-ethylhexyl)phthalate	2.281	1.854	18.7	83	0.00
33	Indeno(1,2,3-cd)pyrene	1.150	1.073	6.7	87	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.091	0.963	11.7	86	0.00
36	Benzo(k)fluoranthene	1.193	1.131	5.2	89	0.00
37 C	Benzo(a)pyrene	1.072	0.974	9.1	87	0.00
38	Dibenzo(a,h)anthracene	0.949	0.919	3.2	93	0.00
39	Benzo(a,h,i)perylene	1.040	0.948	8.8	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011819\
Data File : BE098663.D
Acq On : 18 Jan 2019 8:34
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 18 11:37:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE011519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 15 14:30:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			