

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012020\
 Data File : BE101096.D
 Acq On : 20 Jan 2020 10:05
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 21 02:45:44 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	58	-0.01
2	1,4-Dioxane	1.020	0.970	4.9	59	0.00
3	n-Nitrosodimethylamine	0.556	0.608	-9.4	73	-0.01
4 S	2-Fluorophenol	0.900	1.126	-25.1#	73	-0.01
5 S	Phenol-d6	1.139	1.116	2.0	64	-0.01
6	bis(2-Chloroethyl)ether	1.357	1.414	-4.2	66	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
8 S	Nitrobenzene-d5	0.286	0.260	9.1	61	0.00
9	Naphthalene	4.441	4.298	3.2	57	0.00
10	Hexachlorobutadiene	0.188	0.190	-1.1	59	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.783	-2.6	62	-0.02
12	2-Methylnaphthalene	0.672	0.655	2.5	60	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.113	0.099	12.4	73	0.00
15 S	2-Fluorobiphenyl	1.519	1.368	9.9	64	-0.02
16	Acenaphthylene	1.629	1.485	8.8	72	0.00
17	Acenaphthene	1.166	1.086	6.9	69	-0.02
18	Fluorene	1.474	1.454	1.4	72	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
20	4-Bromophenyl-phenylether	0.195	0.187	4.1	83	-0.01
21	Hexachlorobenzene	0.214	0.193	9.8	73	-0.01
22	Pentachlorophenol	0.053	0.029	45.3#	58	0.00
23	Phenanthrene	1.092	1.014	7.1	77	0.00
24	Anthracene	1.031	0.935	9.3	78	0.00
25 SURR	Fluoranthene-d10	5.555	5.839	-5.1	89	-0.01
26	Fluoranthene	1.368	1.418	-3.7	90	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	124	-0.01
28	Pvrene	1.489	1.125	24.4	91	0.00
29 S	Terphenyl-d14	0.982	0.787	19.9	98	0.00
30	Benzo(a)anthracene	1.141	1.066	6.6	120	0.00
31	Chrysene	1.696	1.671	1.5	118	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.098	-10.3	157#	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.068	20.3	101	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	132	0.00
35	Benzo(b)fluoranthene	1.122	1.060	5.5	134	-0.01
36	Benzo(k)fluoranthene	1.626	1.446	11.1	118	0.00
37 C	Benzo(a)pyrene	1.085	1.035	4.6	131	0.00
38	Dibenzo(a,h)anthracene	0.998	0.761	23.7	109	0.00
39	Benzo(a,h,i)perylene	1.227	0.880	28.3#	92	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012020\
Data File : BE101096.D
Acq On : 20 Jan 2020 10:05
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 21 02:45:44 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			