

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012220\
 Data File : BE101114.D
 Acq On : 22 Jan 2020 13:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 23 04:20:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	-0.01
2	1,4-Dioxane	1.020	0.946	7.3	68	0.00
3	n-Nitrosodimethylamine	0.556	0.492	11.5	70	-0.01
4 S	2-Fluorophenol	0.900	0.932	-3.6	71	-0.01
5 S	Phenol-d6	1.139	1.022	10.3	69	0.00
6	bis(2-Chloroethyl)ether	1.357	1.111	18.1	61	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
8 S	Nitrobenzene-d5	0.286	0.244	14.7	68	0.01
9	Naphthalene	4.441	4.316	2.8	68	0.00
10	Hexachlorobutadiene	0.188	0.203	-8.0	74	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.755	1.0	71	0.00
12	2-Methylnaphthalene	0.672	0.630	6.3	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
14 S	2,4,6-Tribromophenol	0.113	0.048	57.5#	36#	0.00
15 S	2-Fluorobiphenyl	1.519	1.449	4.6	69	-0.01
16	Acenaphthylene	1.629	1.565	3.9	78	0.00
17	Acenaphthene	1.166	1.105	5.2	71	-0.01
18	Fluorene	1.474	1.373	6.9	69	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
20	4-Bromophenyl-phenylether	0.195	0.181	7.2	74	-0.01
21	Hexachlorobenzene	0.214	0.217	-1.4	76	-0.01
22	Pentachlorophenol	0.053	0.000	100.0#	1#	-0.01
23	Phenanthrene	1.092	0.974	10.8	69	0.00
24	Anthracene	1.031	0.968	6.1	75	0.00
25 SURR	Fluoranthene-d10	5.555	5.508	0.8	78	0.00
26	Fluoranthene	1.368	1.313	4.0	78	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
28	Pvrene	1.489	1.291	13.3	78	0.00
29 S	Terphenyl-d14	0.982	0.880	10.4	81	0.00
30	Benzo(a)anthracene	1.141	1.028	9.9	87	0.00
31	Chrysene	1.696	1.757	-3.6	93	0.00
32	Bis(2-ethylhexyl)phthalate	1.902	2.664	-40.1#	149	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.441	-7.5	102	0.01
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.122	0.949	15.4	99	0.00
36	Benzo(k)fluoranthene	1.626	1.340	17.6	90	0.00
37 C	Benzo(a)pyrene	1.085	1.070	1.4	111	0.00
38	Dibenzo(a,h)anthracene	0.998	0.921	7.7	109	0.00
39	Benzo(a,h,i)perylene	1.227	1.044	14.9	90	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE012220\
Data File : BE101114.D
Acq On : 22 Jan 2020 13:17
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 23 04:20:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			