

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE012419\
 Data File : BE098680.D
 Acq On : 24 Jan 2019 14:40
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE012419

Quant Time: Jan 24 15:18:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.815	152	817	0.40	ng	0.00
7) Naphthalene-d8	10.602	136	3659	0.40	ng	# 0.00
13) Acenaphthene-d10	14.471	164	2446	0.40	ng	0.00
19) Phenanthrene-d10	17.229	188	5841	0.40	ng	0.00
27) Chrysene-d12	21.415	240	6344	0.40	ng	0.00
34) Perylene-d12	23.927	264	6081	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.408	112	943	0.39	ng	0.00
5) Phenol-d6	6.998	99	1311	0.37	ng	0.00
8) Nitrobenzene-d5	8.978	82	1235	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.209	152	2594	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.971	330	373	0.34	ng	-0.01
15) 2-Fluorobiphenyl	13.102	172	3949	0.38	ng	0.00
25) Fluoranthene-d10	19.259	212	31025	0.37	ng	0.00
29) Terphenyl-d14	19.862	244	5887	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.312	88	611	0.40	ng	87
3) n-Nitrosodimethylamine	3.646	42	730	0.38	ng	# 13
6) bis(2-Chloroethyl)ether	7.240	93	932	0.39	ng	89
9) Naphthalene	10.654	128	15731	0.38	ng	99
10) Hexachlorobutadiene	10.940	225	1095	0.39	ng	98
12) 2-Methylnaphthalene	12.281	142	2578	0.37	ng	97
16) Acenaphthylene	14.197	152	4688	0.37	ng	99
17) Acenaphthene	14.543	154	2814	0.38	ng	98
18) Fluorene	15.515	166	3819	0.38	ng	99
20) 4-Bromophenyl-phenylether	16.412	248	1323	0.38	ng	# 83
21) Hexachlorobenzene	16.533	284	1510	0.37	ng	95
22) Pentachlorophenol	16.894	266	303	0.29	ng	93
23) Phenanthrene	17.269	178	6084	0.38	ng	99
24) Anthracene	17.363	178	4817	0.37	ng	97
26) Fluoranthene	19.287	202	7643	0.37	ng	99
28) Pyrene	19.655	202	7801	0.39	ng	99
30) Benzo(a)anthracene	21.403	228	6882	0.36	ng	95
31) Chrysene	21.451	228	8009	0.39	ng	98
32) Bis(2-ethylhexyl)phtha...	21.318	149	11848	0.33	ng	100
33) Indeno(1,2,3-cd)pyrene	26.576	276	8129	0.37	ng	99
35) Benzo(b)fluoranthene	23.157	252	7179	0.37	ng	99
36) Benzo(k)fluoranthene	23.207	252	8065	0.38	ng	100
37) Benzo(a)pyrene	23.813	252	6798	0.36	ng	96
38) Dibenzo(a,h)anthracene	26.601	278	6393	0.37	ng	96
39) Benzo(g,h,i)perylene	27.390	276	7121	0.37	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE012419\
Data File : BE098680.D
Acq On : 24 Jan 2019 14:40
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Quant Time: Jan 24 15:18:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 24 14:44:44 2019
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.815 min Scan# 430
 Delta R.T. 0.008 min
 Lab File: BE098680.D
 Acq: 24 Jan 2019 14:40

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE012419

Tgt Ion: 152 Resp: 817




Ion	Ratio	Lower	Upper
152	100		
150	150.9	121.4	182.2
115	68.5	59.1	88.7



#2
 1,4-Dioxane
 Concen: 0.40 ng
 RT: 3.312 min Scan# 39
 Delta R.T. -0.003 min
 Lab File: BE098680.D
 Acq: 24 Jan 2019 14:40

Tgt Ion: 88 Resp: 611

Ion	Ratio	Lower	Upper
88	100		
58	81.5	73.9	110.9
43	44.0	43.7	65.5









#3
n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.646 min Scan# 68
Delta R.T. -0.003 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion: 42 Resp: 730
Ion Ratio Lower Upper
42 100
74 0.0 58.9 88.3#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.408 min Scan# 221
Delta R.T. 0.009 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion: 112 Resp: 943
Ion Ratio Lower Upper
112 100
64 75.4 62.9 94.3
63 188.8 152.2 228.4









#5
Phenol-d6
Concen: 0.37 ng
RT: 6.998 min Scan# 359
Delta R.T. -0.002 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion: 99 Resp: 1311
Ion Ratio Lower Upper
99 100
42 30.0 24.6 37.0
71 48.4 35.1 52.7



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.240 min Scan# 380
Delta R.T. -0.002 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion: 93 Resp: 932
Ion Ratio Lower Upper
93 100
63 309.0 267.3 400.9
95 31.3 26.2 39.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.602 min Scan# 649
Delta R.T. 0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion:	136	Resp:	3659
Ion	Ratio	Lower	Upper
136	100		
137	12.6	9.7	14.5
54	55.8	35.5	53.3#
68	8.2	5.0	7.6#



#8
Nitrobenzene-d5
Concen: 0.35 ng
RT: 8.978 min Scan# 524
Delta R.T. -0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:	82	Resp:	1235
Ion	Ratio	Lower	Upper
82	100		
128	143.4	96.5	144.7
54	269.6	224.7	337.1









#9
Naphthalene
Concen: 0.38 ng
RT: 10.654 min Scan# 653
Delta R.T. 0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:128	Resp:	15731
Ion Ratio	Lower	Upper
128	100	
129	3.3	2.6 3.8
127	4.2	3.0 4.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.940 min Scan# 675
Delta R.T. 0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:225	Resp:	1095
Ion Ratio	Lower	Upper
225	100	
223	60.5	47.8 71.8
227	65.6	50.5 75.7









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.209 min Scan# 763
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion:152 Resp: 2594
Ion Ratio Lower Upper
152 100
151 19.4 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.281 min Scan# 768
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:142 Resp: 2578
Ion Ratio Lower Upper
142 100
141 89.0 70.0 105.0
115 47.8 35.2 52.8





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.471 min Scan# 920

Delta R.T. 0.000 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40





Instrument :

BNA_E

ClientSampleId :

ICVBE012419

Tgt Ion:	164	Resp:	2446
Ion	Ratio	Lower	Upper
164	100		
162	102.2	80.5	120.7
160	49.6	40.1	60.1



#14

2,4,6-Tribromophenol

Concen: 0.34 ng




RT: 15.971 min Scan# 1026


Delta R.T. -0.011 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40

Tgt Ion:	330	Resp:	373
Ion	Ratio	Lower	Upper
330	100		
332	105.9	73.3	109.9
141	44.2	33.1	49.7









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 13.102 min Scan# 825
Delta R.T. -0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:	172	Resp:	3949
Ion Ratio	Lower	Upper	
172	100		
171	35.1	29.0	43.4
170	22.8	20.8	31.2



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.197 min Scan# 901
Delta R.T. -0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:	152	Resp:	4688
Ion Ratio	Lower	Upper	
152	100		
151	19.5	16.3	24.5
153	13.1	10.7	16.1





#17

Acenaphthene

Concen: 0.38 ng

RT: 14.543 min Scan# 925

Delta R.T. 0.000 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40





Instrument :

BNA_E

ClientSampleId :

ICVBE012419

Tgt Ion:	154	Resp:	2814
Ion	Ratio	Lower	Upper
154	100		
153	115.5	93.9	140.9
152	56.1	45.8	68.8



#18

Fluorene

Concen: 0.38 ng




RT: 15.515 min Scan# 992


Delta R.T. 0.002 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40

Tgt Ion:	166	Resp:	3819
Ion	Ratio	Lower	Upper
166	100		
165	100.2	79.3	118.9
167	14.3	10.6	15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.229 min Scan# 1120
Delta R.T. 0.002 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion:188 Resp: 5841
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.6 4.7 7.1



#20
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.412 min Scan# 1059
Delta R.T. -0.012 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:248 Resp: 1323
Ion Ratio Lower Upper
248 100
250 88.7 80.2 120.4
141 95.3 60.5 90.7#









#21
Hexachlorobenzene
Concen: 0.37 ng
RT: 16.533 min Scan# 1068
Delta R.T. 0.002 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion:284 Resp: 1510
Ion Ratio Lower Upper
284 100
142 35.4 25.5 38.3
249 35.8 26.6 39.8



#22
Pentachlorophenol
Concen: 0.29 ng
RT: 16.894 min Scan# 1095
Delta R.T. 0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:266 Resp: 303
Ion Ratio Lower Upper
266 100
264 77.9 53.7 80.5
268 75.3 59.8 89.8









#23
Phenanthrene
Concen: 0.38 ng
RT: 17.269 min Scan# 1123
Delta R.T. 0.001 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:178 Resp: 6084
Ion Ratio Lower Upper
178 100
176 19.5 15.9 23.9
179 15.0 12.2 18.4



#24
Anthracene
Concen: 0.37 ng
RT: 17.363 min Scan# 1130
Delta R.T. 0.002 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:178 Resp: 4817
Ion Ratio Lower Upper
178 100
176 19.9 14.9 22.3
179 16.0 11.8 17.8









#25
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.259 min Scan# 1391
Delta R.T. -0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419


Tgt Ion:212 Resp: 31025
Ion Ratio Lower Upper
212 100
106 1.9 1.7 2.5
104 1.1 1.0 1.6



#26
Fluoranthene
Concen: 0.37 ng
RT: 19.287 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:202 Resp: 7643
Ion Ratio Lower Upper
202 100
101 7.7 6.7 10.1
203 16.9 13.3 19.9





#27

Chrysene-d12

Concen: 0.40 ng

RT: 21.415 min Scan# 1634

Delta R.T. -0.000 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40





Tgt Ion:240 Resp: 6344

Ion Ratio Lower Upper

240 100

120 6.9 4.7 7.1

236 26.8 21.9 32.9



#28

Pyrene

Concen: 0.39 ng

RT: 19.655 min Scan# 1460

Delta R.T. 0.006 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40




Tgt Ion:202 Resp: 7801

Ion Ratio Lower Upper

202 100

200 21.2 17.0 25.6

203 16.9 13.8 20.8





#29

Terphenyl-d14

Concen: 0.37 ng

RT: 19.862 min Scan# 1496

Delta R.T. 0.006 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40





Tgt Ion:244 Resp: 5887

Ion Ratio Lower Upper

244 100

212 39.1 29.4 44.2

122 8.1 7.1 10.7



#30

Benzo(a)anthracene

Concen: 0.36 ng

RT: 21.403 min Scan# 1633

Delta R.T. 0.012 min

Lab File: BE098680.D

Acq: 24 Jan 2019 14:40




Tgt Ion:228 Resp: 6882


Ion Ratio Lower Upper

228 100

226 27.1 23.0 34.4

229 21.8 15.0 22.6









#31
Chrysene
Concen: 0.39 ng
RT: 21.451 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:	228	Resp:	8009
Ion Ratio	Lower	Upper	
228	100		
226	29.7	24.6	36.8
229	19.9	16.4	24.6



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.33 ng
RT: 21.318 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:	149	Resp:	11848
Ion Ratio	Lower	Upper	
149	100		
167	6.9	5.4	8.2
279	1.4	1.0	1.4









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 26.576 min Scan# 2955
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:276 Resp: 8129
Ion Ratio Lower Upper
276 100
138 15.9 13.3 19.9
277 24.1 19.4 29.0



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.927 min Scan# 2216
Delta R.T. -0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:264 Resp: 6081
Ion Ratio Lower Upper
264 100
260 25.1 20.5 30.7
265 28.2 22.3 33.5









#35
Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 23.157 min Scan# 2000
Delta R.T. 0.003 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:252 Resp: 7179
Ion Ratio Lower Upper
252 100
253 24.2 19.4 29.0
125 8.6 8.2 12.4



#36
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 23.207 min Scan# 2014
Delta R.T. 0.003 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:252 Resp: 8065
Ion Ratio Lower Upper
252 100
253 23.9 19.0 28.4
125 8.5 7.0 10.4









#37
Benzo(a)pyrene
Concen: 0.36 ng
RT: 23.813 min Scan# 2184
Delta R.T. -0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40




Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:252 Resp: 6798
Ion Ratio Lower Upper
252 100
253 24.4 21.0 31.4
125 9.8 9.2 13.8



#38
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 26.601 min Scan# 2962
Delta R.T. 0.004 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Tgt Ion:278 Resp: 6393
Ion Ratio Lower Upper
278 100
139 13.9 12.7 19.1
279 25.8 22.0 33.0





#39
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 27.390 min Scan# 3183
Delta R.T. 0.000 min
Lab File: BE098680.D
Acq: 24 Jan 2019 14:40

Instrument :
BNA_E
ClientSampleId :
ICVBE012419

Tgt Ion:	276	Resp:	7121
Ion	Ratio	Lower	Upper
276	100		
277	26.1	20.3	30.5
138	16.2	13.3	19.9

