

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE012419\
 Data File : BE098680.D
 Acq On : 24 Jan 2019 15:28
 Operator : SJ/JU
 Sample : SSTDIC0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE012419

Quant Time: Jan 24 15:59:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.01
2	1,4-Dioxane	0.400	0.400	0.0	107	-0.01
3	n-Nitrosodimethylamine	0.400	0.399	0.3	104	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	96	0.01
5 S	Phenol-d6	0.400	0.365	8.8	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.341	14.8	93	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.01
8 S	Nitrobenzene-d5	0.400	0.370	7.5	93	0.01
9	Naphthalene	0.400	0.375	6.3	98	0.01
10	Hexachlorobutadiene	0.400	0.393	1.8	102	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.375	6.3	96	0.02
12	2-Methylnaphthalene	0.400	0.379	5.3	98	0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.02
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	100	0.00
16	Acenaphthylene	0.400	0.370	7.5	98	0.00
17	Acenaphthene	0.400	0.364	9.0	93	0.00
18	Fluorene	0.400	0.374	6.5	99	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.400	0.351	12.3	94	0.00
21	Hexachlorobenzene	0.400	0.375	6.3	98	0.00
22	Pentachlorophenol	0.400	0.431	-7.7	124	0.00
23	Phenanthrene	0.400	0.369	7.8	99	0.00
24	Anthracene	0.400	0.359	10.3	96	0.00
25 SURR	Fluoranthene-d10	0.400	0.362	9.5	96	0.00
26	Fluoranthene	0.400	0.360	10.0	96	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
28	Pyrene	0.400	0.385	3.8	97	0.00
29 S	Terphenyl-d14	0.400	0.370	7.5	92	0.00
30	Benzo(a)anthracene	0.400	0.353	11.8	92	0.01
31	Chrysene	0.400	0.388	3.0	98	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.324	19.0	83	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	101	0.00
34 I	Perylene-d12	0.400	0.400	0.0	100	0.00
35	Benzo(b)fluoranthene	0.400	0.364	9.0	92	0.00
36	Benzo(k)fluoranthene	0.400	0.371	7.3	96	0.00
37 C	Benzo(a)pyrene	0.400	0.366	8.5	96	0.00
38	Dibenzo(a,h)anthracene	0.400	0.365	8.8	100	0.00
39	Benzo(g,h,i)perylene	0.400	0.371	7.3	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0