

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE012419\
 Data File : BE098680.D
 Acq On : 24 Jan 2019 15:28
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE012419

Quant Time: Jan 24 15:59:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.01
2	1,4-Dioxane	0.739	0.746	-0.9	107	-0.01
3	n-Nitrosodimethylamine	0.877	0.942	-7.4	104	0.00
4 S	2-Fluorophenol	1.182	1.150	2.7	96	0.01
5 S	Phenol-d6	1.737	1.584	8.8	96	0.00
6	bis(2-Chloroethyl)ether	1.183	1.008	14.8	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.01
8 S	Nitrobenzene-d5	0.389	0.357	8.2	93	0.01
9	Naphthalene	4.584	4.301	6.2	98	0.01
10	Hexachlorobutadiene	0.307	0.301	2.0	102	0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.717	6.3	96	0.02
12	2-Methylnaphthalene	0.761	0.722	5.1	98	0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.02
14 S	2,4,6-Tribromophenol	0.177	0.159	10.2	92	0.00
15 S	2-Fluorobiphenyl	1.714	1.619	5.5	100	0.00
16	Acenaphthylene	2.052	1.900	7.4	98	0.00
17	Acenaphthene	1.212	1.103	9.0	93	0.00
18	Fluorene	1.656	1.549	6.5	99	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.236	0.207	12.3	94	0.00
21	Hexachlorobenzene	0.278	0.260	6.5	98	0.00
22	Pentachlorophenol	0.071	0.077	-8.5	124	0.00
23	Phenanthrene	1.101	1.014	7.9	99	0.00
24	Anthracene	0.919	0.807	12.2	96	0.00
25 SURR	Fluoranthene-d10	5.713	5.176	9.4	96	0.00
26	Fluoranthene	1.417	1.277	9.9	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
28	Pyrene	1.274	1.227	3.7	97	0.00
29 S	Terphenyl-d14	0.998	0.924	7.4	92	0.00
30	Benzo(a)anthracene	1.207	1.064	11.8	92	0.01
31	Chrysene	1.308	1.268	3.1	98	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	1.813	19.0	83	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.295	5.5	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.294	1.176	9.1	92	0.00
36	Benzo(k)fluoranthene	1.383	1.283	7.2	96	0.00
37 C	Benzo(a)pyrene	1.237	1.132	8.5	96	0.00
38	Dibenzo(a,h)anthracene	1.146	1.047	8.6	100	0.00
39	Benzo(g,h,i)perylene	1.258	1.166	7.3	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0