

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE012419\
 Data File : BE098674.D
 Acq On : 24 Jan 2019 10:57
 Operator : SJ/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jan 24 13:06:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 12:37:26 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	783	0.40	ng	0.00
7) Naphthalene-d8	10.614	136	3592	0.40	ng	0.01
13) Acenaphthene-d10	14.470	164	2395	0.40	ng	0.00
19) Phenanthrene-d10	17.228	188	5789	0.40	ng	# 0.00
27) Chrysene-d12	21.414	240	6298	0.40	ng	0.00
34) Perylene-d12	23.926	264	6114	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	448	0.21	ng	0.00
5) Phenol-d6	7.010	99	676	0.23	ng	0.00
8) Nitrobenzene-d5	8.990	82	604	0.21	ng	0.01
11) 2-Methylnaphthalene-d10	12.209	152	1324	0.22	ng	0.00
14) 2,4,6-Tribromophenol	15.983	330	178	0.17	ng	0.00
15) 2-Fluorobiphenyl	13.102	172	1892	0.19	ng	0.00
25) Fluoranthene-d10	19.265	212	15502	0.22	ng	0.00
29) Terphenyl-d14	19.862	244	2970	0.21	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.313	88	352	0.23	ng	85
3) n-Nitrosodimethylamine	3.659	42	163	0.14	ng	# 13
6) bis(2-Chloroethyl)ether	7.252	93	492	0.22	ng	# 1
9) Naphthalene	10.653	128	7929	0.22	ng	99
10) Hexachlorobutadiene	10.939	225	538	0.21	ng	95
12) 2-Methylnaphthalene	12.295	142	1312	0.22	ng	96
16) Acenaphthylene	14.197	152	2252	0.22	ng	100
17) Acenaphthene	14.543	154	1401	0.22	ng	99
18) Fluorene	15.528	166	1883	0.22	ng	100
20) 4-Bromophenyl-phenylether	16.425	248	619	0.20	ng	# 92
21) Hexachlorobenzene	16.532	284	761	0.21	ng	96
22) Pentachlorophenol	16.907	266	201	0.19	ng	86
23) Phenanthrene	17.268	178	3021	0.22	ng	99
24) Anthracene	17.362	178	2205	0.19	ng	98
26) Fluoranthene	19.294	202	3817	0.22	ng	98
28) Pyrene	19.656	202	3875	0.22	ng	99
30) Benzo(a)anthracene	21.402	228	3412	0.20	ng	96
31) Chrysene	21.450	228	3929	0.22	ng	97
32) Bis(2-ethylhexyl)phtha...	21.317	149	6745	0.19	ng	# 99
33) Indeno(1,2,3-cd)pyrene	26.580	276	3740	0.21	ng	98
35) Benzo(b)fluoranthene	23.160	252	3596	0.22	ng	95
36) Benzo(k)fluoranthene	23.210	252	4020	0.22	ng	# 93
37) Benzo(a)pyrene	23.819	252	3511	0.21	ng	# 90
38) Dibenzo(a,h)anthracene	26.602	278	2720	0.19	ng	94
39) Benzo(g,h,i)perylene	27.394	276	3501	0.22	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.816 min Scan# 430
 Delta R.T. 0.009 min
 Lab File: BE098674.D
 Acq: 24 Jan 2019 10:57

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.2



Tgt Ion:152 Resp: 783




Ion	Ratio	Lower	Upper
152	100		
150	140.2	121.4	182.2
115	69.8	59.1	88.7




#2
 1,4-Dioxane
 Concen: 0.23 ng
 RT: 3.313 min Scan# 39
 Delta R.T. -0.002 min
 Lab File: BE098674.D
 Acq: 24 Jan 2019 10:57

Tgt Ion: 88 Resp: 352

Ion	Ratio	Lower	Upper
88	100		
58	76.1	73.9	110.9
43	47.2	43.7	65.5









#3
n-Nitrosodimethylamine
Concen: 0.14 ng
RT: 3.659 min Scan# 69
Delta R.T. 0.010 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 163
Ion Ratio Lower Upper
42 100
74 0.0 58.9 88.3#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.21 ng
RT: 5.409 min Scan# 221
Delta R.T. 0.010 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion: 112 Resp: 448
Ion Ratio Lower Upper
112 100
64 87.5 62.9 94.3
63 198.0 152.2 228.4









#5
Phenol-d6
Concen: 0.23 ng
RT: 7.010 min Scan# 360
Delta R.T. 0.010 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion: 99 Resp: 676
Ion Ratio Lower Upper
99 100
42 38.3 24.6 37.0#
71 41.4 35.1 52.7



#6
bis(2-Chloroethyl)ether
Concen: 0.22 ng
RT: 7.252 min Scan# 381
Delta R.T. 0.010 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion: 93 Resp: 492
Ion Ratio Lower Upper
93 100
63 0.0 267.3 400.9#
95 25.4 26.2 39.4#









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.614 min Scan# 650
Delta R.T. 0.012 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	3592
Ion Ratio	Lower	Upper	
136	100		
137	14.1	9.7	14.5
54	40.1	35.5	53.3
68	6.6	5.0	7.6



#8
Nitrobenzene-d5
Concen: 0.21 ng
RT: 8.990 min Scan# 525
Delta R.T. 0.012 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	82	Resp:	604
Ion Ratio	Lower	Upper	
82	100		
128	196.9	96.5	144.7#
54	319.8	224.7	337.1









#9
Naphthalene
Concen: 0.22 ng
RT: 10.653 min Scan# 653
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:	128	Resp:	7929
Ion	Ratio	Lower	Upper
128	100		
129	3.7	2.6	3.8
127	4.3	3.0	4.6



#10
Hexachlorobutadiene
Concen: 0.21 ng
RT: 10.939 min Scan# 675
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	225	Resp:	538
Ion	Ratio	Lower	Upper
225	100		
223	67.3	47.8	71.8
227	63.2	50.5	75.7









#11
2-Methylnaphthalene-d10
Concen: 0.22 ng
RT: 12.209 min Scan# 763
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:152 Resp: 1324
Ion Ratio Lower Upper
152 100
151 18.8 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.22 ng
RT: 12.295 min Scan# 769
Delta R.T. 0.014 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:142 Resp: 1312
Ion Ratio Lower Upper
142 100
141 85.7 70.0 105.0
115 48.9 35.2 52.8









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.470 min Scan# 920
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:164 Resp: 2395
Ion Ratio Lower Upper
164 100
162 104.3 80.5 120.7
160 48.4 40.1 60.1



#14
2,4,6-Tribromophenol
Concen: 0.17 ng
RT: 15.983 min Scan# 1027
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:330 Resp: 178
Ion Ratio Lower Upper
330 100
332 97.8 73.3 109.9
141 39.3 33.1 49.7









#15
2-Fluorobiphenyl
Concen: 0.19 ng
RT: 13.102 min Scan# 825
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	1892
Ion Ratio	Lower	Upper	
172	100		
171	36.4	29.0	43.4
170	25.9	20.8	31.2



#16
Acenaphthylene
Concen: 0.22 ng
RT: 14.197 min Scan# 901
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	152	Resp:	2252
Ion Ratio	Lower	Upper	
152	100		
151	20.4	16.3	24.5
153	13.1	10.7	16.1









#17
Acenaphthene
Concen: 0.22 ng
RT: 14.543 min Scan# 925
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:	154	Resp:	1401
Ion	Ratio	Lower	Upper
154	100		
153	117.5	93.9	140.9
152	58.4	45.8	68.8



#18
Fluorene
Concen: 0.22 ng
RT: 15.528 min Scan# 993
Delta R.T. 0.014 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	166	Resp:	1883
Ion	Ratio	Lower	Upper
166	100		
165	99.2	79.3	118.9
167	14.4	10.6	15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.228 min Scan# 1120
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:188 Resp: 5789
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.0 4.7 7.1#



#20
4-Bromophenyl-phenylether
Concen: 0.20 ng
RT: 16.425 min Scan# 1060
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:248 Resp: 619
Ion Ratio Lower Upper
248 100
250 99.7 80.2 120.4
141 60.4 60.5 90.7#









#21
Hexachlorobenzene
Concen: 0.21 ng
RT: 16.532 min Scan# 1068
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:	284	Resp:	761
Ion Ratio	Lower	Upper	
284	100		
142	36.5	25.5	38.3
249	33.5	26.6	39.8



#22
Pentachlorophenol
Concen: 0.19 ng
RT: 16.907 min Scan# 1096
Delta R.T. 0.014 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	266	Resp:	201
Ion Ratio	Lower	Upper	
266	100		
264	78.8	53.7	80.5
268	86.5	59.8	89.8









#23
Phenanthrene
Concen: 0.22 ng
RT: 17.268 min Scan# 1123
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:178 Resp: 3021
Ion Ratio Lower Upper
178 100
176 19.8 15.9 23.9
179 14.5 12.2 18.4



#24
Anthracene
Concen: 0.19 ng
RT: 17.362 min Scan# 1130
Delta R.T. 0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:178 Resp: 2205
Ion Ratio Lower Upper
178 100
176 17.8 14.9 22.3
179 15.8 11.8 17.8









#25
Fluoranthene-d10
Concen: 0.22 ng
RT: 19.265 min Scan# 1392
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:212 Resp: 15502
Ion Ratio Lower Upper
212 100
106 2.0 1.7 2.5
104 1.1 1.0 1.6



#26
Fluoranthene
Concen: 0.22 ng
RT: 19.294 min Scan# 1397
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:202 Resp: 3817
Ion Ratio Lower Upper
202 100
101 7.7 6.7 10.1
203 17.2 13.3 19.9









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.414 min Scan# 1634
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:240 Resp: 6298
Ion Ratio Lower Upper
240 100
120 6.7 4.7 7.1
236 30.0 21.9 32.9



#28
Pyrene
Concen: 0.22 ng
RT: 19.656 min Scan# 1460
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:202 Resp: 3875
Ion Ratio Lower Upper
202 100
200 21.2 17.0 25.6
203 18.3 13.8 20.8









#29
Terphenyl-d14
Concen: 0.21 ng
RT: 19.862 min Scan# 1496
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:244 Resp: 2970
Ion Ratio Lower Upper
244 100
212 45.2 29.4 44.2#
122 10.4 7.1 10.7



#30
Benzo(a)anthracene
Concen: 0.20 ng
RT: 21.402 min Scan# 1633
Delta R.T. 0.011 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:228 Resp: 3412
Ion Ratio Lower Upper
228 100
226 29.0 23.0 34.4
229 22.5 15.0 22.6









#31
Chrysene
Concen: 0.22 ng
RT: 21.450 min Scan# 1637
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:	228	Resp:	3929
Ion Ratio	Lower	Upper	
228	100		
226	33.1	24.6	36.8
229	20.7	16.4	24.6



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.19 ng
RT: 21.317 min Scan# 1626
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:	149	Resp:	6745
Ion Ratio	Lower	Upper	
149	100		
167	6.6	5.4	8.2
279	1.6	1.0	1.4









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.21 ng
RT: 26.580 min Scan# 2956
Delta R.T. 0.004 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:276 Resp: 3740
Ion Ratio Lower Upper
276 100
138 15.4 13.3 19.9
277 23.9 19.4 29.0



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.926 min Scan# 2216
Delta R.T. -0.001 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:264 Resp: 6114
Ion Ratio Lower Upper
264 100
260 25.9 20.5 30.7
265 28.8 22.3 33.5









#35
Benzo(b)fluoranthene
Concen: 0.22 ng
RT: 23.160 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.2


Tgt Ion:252 Resp: 3596
Ion Ratio Lower Upper
252 100
253 27.2 19.4 29.0
125 10.8 8.2 12.4



#36
Benzo(k)fluoranthene
Concen: 0.22 ng
RT: 23.210 min Scan# 2015
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Tgt Ion:252 Resp: 4020
Ion Ratio Lower Upper
252 100
253 27.2 19.0 28.4
125 11.3 7.0 10.4#









#37
Benzo(a)pyrene
Concen: 0.21 ng
RT: 23.819 min Scan# 2186
Delta R.T. 0.006 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 3511
Ion Ratio Lower Upper
252 100
253 32.3 21.0 31.4#
125 13.9 9.2 13.8#



#38
Dibenzo(a,h)anthracene
Concen: 0.19 ng
RT: 26.602 min Scan# 2962
Delta R.T. 0.004 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57


Tgt Ion:278 Resp: 2720
Ion Ratio Lower Upper
278 100
139 18.8 12.7 19.1
279 30.1 22.0 33.0





#39
Benzo(g,h,i)perylene
Concen: 0.22 ng
RT: 27.394 min Scan# 3184
Delta R.T. 0.004 min
Lab File: BE098674.D
Acq: 24 Jan 2019 10:57

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.2



Tgt Ion:	276	Resp:	3501
Ion	Ratio	Lower	Upper
276	100		
277	29.8	20.3	30.5
138	17.6	13.3	19.9

