

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020119\
 Data File : BE098686.D
 Acq On : 1 Feb 2019 10:37
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 01 14:32:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.739	0.778	-5.3	117	0.00
3	n-Nitrosodimethylamine	0.877	0.829	5.5	96	0.00
4 S	2-Fluorophenol	1.182	1.109	6.2	97	0.00
5 S	Phenol-d6	1.737	1.540	11.3	98	0.00
6	bis(2-Chloroethyl)ether	1.183	1.043	11.8	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.389	0.347	10.8	93	0.00
9	Naphthalene	4.584	4.504	1.7	106	0.00
10	Hexachlorobutadiene	0.307	0.305	0.7	106	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.747	2.4	103	0.00
12	2-Methylnaphthalene	0.761	0.722	5.1	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.177	0.158	10.7	94	-0.01
15 S	2-Fluorobiphenyl	1.714	1.628	5.0	103	-0.02
16	Acenaphthylene	2.052	1.910	6.9	102	0.00
17	Acenaphthene	1.212	1.186	2.1	103	-0.01
18	Fluorene	1.656	1.573	5.0	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	-0.01
20	4-Bromophenyl-phenylether	0.236	0.220	6.8	100	-0.01
21	Hexachlorobenzene	0.278	0.281	-1.1	106	0.00
22	Pentachlorophenol	0.071	0.062	12.7	99	0.00
23	Phenanthrene	1.101	1.055	4.2	102	0.00
24	Anthracene	0.919	0.842	8.4	100	0.00
25 SURR	Fluoranthene-d10	5.713	5.437	4.8	101	0.00
26	Fluoranthene	1.417	1.343	5.2	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	1.274	1.269	0.4	102	0.00
29 S	Terphenyl-d14	0.998	0.962	3.6	97	0.00
30	Benzo(a)anthracene	1.207	1.163	3.6	103	0.00
31	Chrysene	1.308	1.273	2.7	100	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	1.847	17.5	86	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.236	9.8	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.294	1.251	3.3	98	0.00
36	Benzo(k)fluoranthene	1.383	1.373	0.7	103	0.00
37 C	Benzo(a)pyrene	1.237	1.202	2.8	101	0.00
38	Dibenzo(a,h)anthracene	1.146	0.961	16.1	92	-0.01
39	Benzo(g,h,i)perylene	1.258	1.173	6.8	98	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0