

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020119\
 Data File : BE098698.D
 Acq On : 1 Feb 2019 19:22
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Feb 02 03:15:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	895	0.40	ng	0.00
7) Naphthalene-d8	10.601	136	4034	0.40	ng	0.00
13) Acenaphthene-d10	14.470	164	2700	0.40	ng	0.00
19) Phenanthrene-d10	17.228	188	6435	0.40	ng	0.00
27) Chrysene-d12	21.403	240	7168	0.40	ng	#-0.01
34) Perylene-d12	23.923	264	7006	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.397	112	1027	0.39	ng	0.00
5) Phenol-d6	6.998	99	1464	0.38	ng	0.00
8) Nitrobenzene-d5	8.977	82	1479	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.208	152	2993	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.970	330	501	0.42	ng	-0.01
15) 2-Fluorobiphenyl	13.087	172	4586	0.40	ng	-0.02
25) Fluoranthene-d10	19.259	212	34703	0.38	ng	0.00
29) Terphenyl-d14	19.857	244	7006	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.312	88	657	0.39	ng	98
3) n-Nitrosodimethylamine	3.646	42	773	0.37	ng	# 89
6) bis(2-Chloroethyl)ether	7.240	93	983	0.37	ng	100
9) Naphthalene	10.653	128	18249	0.39	ng	100
10) Hexachlorobutadiene	10.926	225	1199	0.39	ng	95
12) 2-Methylnaphthalene	12.280	142	2946	0.38	ng	95
16) Acenaphthylene	14.196	152	5318	0.38	ng	98
17) Acenaphthene	14.542	154	3243	0.40	ng	97
18) Fluorene	15.515	166	4439	0.40	ng	99
20) 4-Bromophenyl-phenylether	16.411	248	1489	0.39	ng	93
21) Hexachlorobenzene	16.532	284	1760	0.39	ng	98
22) Pentachlorophenol	16.893	266	400	0.35	ng	99
23) Phenanthrene	17.268	178	6923	0.39	ng	100
24) Anthracene	17.362	178	5529	0.38	ng	98
26) Fluoranthene	19.288	202	8788	0.39	ng	98
28) Pyrene	19.650	202	8987	0.39	ng	100
30) Benzo(a)anthracene	21.391	228	8638	0.40	ng	97
31) Chrysene	21.451	228	8839	0.38	ng	99
32) Bis(2-ethylhexyl)phtha...	21.318	149	14369	0.36	ng	99
33) Indeno(1,2,3-cd)pyrene	26.566	276	9709	0.40	ng	# 92
35) Benzo(b)fluoranthene	23.149	252	8670	0.38	ng	99
36) Benzo(k)fluoranthene	23.199	252	9253	0.38	ng	99
37) Benzo(a)pyrene	23.809	252	8089	0.37	ng	98
38) Dibenzo(a,h)anthracene	26.595	278	7655	0.38	ng	98
39) Benzo(g,h,i)perylene	27.376	276	8212	0.37	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020119\
Data File : BE098698.D
Acq On : 1 Feb 2019 19:22
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Feb 02 03:15:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 24 14:44:44 2019
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.804 min Scan# 429
 Delta R.T. -0.003 min
 Lab File: BE098698.D
 Acq: 1 Feb 2019 19:22

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 152 Resp: 895




Ion	Ratio	Lower	Upper
152	100		
150	150.9	121.4	182.2
115	73.2	59.1	88.7




#2
 1,4-Dioxane
 Concen: 0.39 ng
 RT: 3.312 min Scan# 39
 Delta R.T. -0.003 min
 Lab File: BE098698.D
 Acq: 1 Feb 2019 19:22

Tgt Ion: 88 Resp: 657

Ion	Ratio	Lower	Upper
88	100		
58	91.3	73.9	110.9
43	51.3	43.7	65.5









#3
n-Nitrosodimethylamine
Concen: 0.37 ng
RT: 3.646 min Scan# 68
Delta R.T. -0.003 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 773
Ion Ratio Lower Upper
42 100
74 83.1 58.9 88.3
44 3.6 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.397 min Scan# 220
Delta R.T. -0.002 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion: 112 Resp: 1027
Ion Ratio Lower Upper
112 100
64 78.1 62.9 94.3
63 197.0 152.2 228.4









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.998 min Scan# 359
Delta R.T. -0.002 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 1464
Ion Ratio Lower Upper
99 100
42 28.3 24.6 37.0
71 48.7 35.1 52.7



#6
bis(2-Chloroethyl)ether
Concen: 0.37 ng
RT: 7.240 min Scan# 380
Delta R.T. -0.002 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion: 93 Resp: 983
Ion Ratio Lower Upper
93 100
63 334.1 267.3 400.9
95 33.9 26.2 39.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.601 min Scan# 649
Delta R.T. -0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	4034
Ion	Ratio	Lower	Upper
136	100		
137	13.1	9.7	14.5
54	46.6	35.5	53.3
68	7.4	5.0	7.6



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.977 min Scan# 524
Delta R.T. -0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:	82	Resp:	1479
Ion	Ratio	Lower	Upper
82	100		
128	159.1	96.5	144.7#
54	257.2	224.7	337.1









#9
Naphthalene
Concen: 0.39 ng
RT: 10.653 min Scan# 653
Delta R.T. -0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128	Resp:	18249
Ion Ratio	Lower	Upper
128	100	
129	3.3	2.6 3.8
127	3.9	3.0 4.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.926 min Scan# 674
Delta R.T. -0.013 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:225	Resp:	1199
Ion Ratio	Lower	Upper
225	100	
223	65.1	47.8 71.8
227	64.8	50.5 75.7









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.208 min Scan# 763
Delta R.T. -0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 2993
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.280 min Scan# 768
Delta R.T. -0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:142 Resp: 2946
Ion Ratio Lower Upper
142 100
141 91.0 70.0 105.0
115 48.5 35.2 52.8





#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.470 min Scan# 920

Delta R.T. -0.001 min

Lab File: BE098698.D

Acq: 1 Feb 2019 19:22





Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	164	Resp:	2700
Ion	Ratio	Lower	Upper
164	100		
162	102.8	80.5	120.7
160	49.9	40.1	60.1



#14

2,4,6-Tribromophenol

Concen: 0.42 ng




RT: 15.970 min Scan# 1026


Delta R.T. -0.012 min

Lab File: BE098698.D

Acq: 1 Feb 2019 19:22

Tgt Ion:	330	Resp:	501
Ion	Ratio	Lower	Upper
330	100		
332	95.2	73.3	109.9
141	39.5	33.1	49.7









#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 13.087 min Scan# 824
Delta R.T. -0.016 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172	Resp:	4586
Ion Ratio	Lower	Upper
172	100	
171	35.1	29.0 43.4
170	25.4	20.8 31.2



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.196 min Scan# 901
Delta R.T. -0.002 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:152	Resp:	5318
Ion Ratio	Lower	Upper
152	100	
151	19.6	16.3 24.5
153	12.6	10.7 16.1









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.542 min Scan# 925
Delta R.T. -0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 3243
Ion Ratio Lower Upper
154 100
153 113.9 93.9 140.9
152 54.7 45.8 68.8



#18
Fluorene
Concen: 0.40 ng
RT: 15.515 min Scan# 992
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:166 Resp: 4439
Ion Ratio Lower Upper
166 100
165 99.6 79.3 118.9
167 13.5 10.6 15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.228 min Scan# 1120
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 6435
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.0 4.7 7.1



#20
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.411 min Scan# 1059
Delta R.T. -0.013 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:248 Resp: 1489
Ion Ratio Lower Upper
248 100
250 94.1 80.2 120.4
141 68.8 60.5 90.7









#21
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.532 min Scan# 1068
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:284 Resp: 1760
Ion Ratio Lower Upper
284 100
142 33.8 25.5 38.3
249 32.7 26.6 39.8



#22
Pentachlorophenol
Concen: 0.35 ng
RT: 16.893 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:266 Resp: 400
Ion Ratio Lower Upper
266 100
264 67.3 53.7 80.5
268 76.2 59.8 89.8









#23
Phenanthrene
Concen: 0.39 ng
RT: 17.268 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 6923
Ion Ratio Lower Upper
178 100
176 20.0 15.9 23.9
179 15.1 12.2 18.4



#24
Anthracene
Concen: 0.38 ng
RT: 17.362 min Scan# 1130
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:178 Resp: 5529
Ion Ratio Lower Upper
178 100
176 19.9 14.9 22.3
179 15.6 11.8 17.8









#25
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.259 min Scan# 1391
Delta R.T. 0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 34703
Ion Ratio Lower Upper
212 100
106 2.0 1.7 2.5
104 1.2 1.0 1.6



#26
Fluoranthene
Concen: 0.39 ng
RT: 19.288 min Scan# 1396
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:202 Resp: 8788
Ion Ratio Lower Upper
202 100
101 7.6 6.7 10.1
203 17.2 13.3 19.9









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.403 min Scan# 1633
Delta R.T. -0.012 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 7168
Ion Ratio Lower Upper
240 100
120 10.1 4.7 7.1#
236 29.1 21.9 32.9



#28
Pyrene
Concen: 0.39 ng
RT: 19.650 min Scan# 1459
Delta R.T. 0.001 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:202 Resp: 8987
Ion Ratio Lower Upper
202 100
200 21.2 17.0 25.6
203 17.4 13.8 20.8





#29

Terphenyl-d14

Concen: 0.39 ng

RT: 19.857 min Scan# 1495

Delta R.T. 0.001 min

Lab File: BE098698.D

Acq: 1 Feb 2019 19:22





Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:244	Resp:	7006
Ion Ratio	Lower	Upper
244	100	
212	36.7	29.4 44.2
122	8.6	7.1 10.7



#30

Benzo(a)anthracene

Concen: 0.40 ng




RT: 21.391 min Scan# 1632


Delta R.T. -0.000 min

Lab File: BE098698.D

Acq: 1 Feb 2019 19:22

Tgt Ion:228	Resp:	8638
Ion Ratio	Lower	Upper
228	100	
226	28.0	23.0 34.4
229	21.0	15.0 22.6









#31
Chrysene
Concen: 0.38 ng
RT: 21.451 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	8839
Ion	Ratio	Lower	Upper
228	100		
226	31.4	24.6	36.8
229	21.2	16.4	24.6



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.36 ng
RT: 21.318 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:	149	Resp:	14369
Ion	Ratio	Lower	Upper
149	100		
167	7.1	5.4	8.2
279	1.3	1.0	1.4









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.40 ng
RT: 26.566 min Scan# 2952
Delta R.T. -0.010 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 9709
Ion Ratio Lower Upper
276 100
138 8.1 13.3 19.9#
277 24.0 19.4 29.0



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.923 min Scan# 2215
Delta R.T. -0.004 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:264 Resp: 7006
Ion Ratio Lower Upper
264 100
260 25.6 20.5 30.7
265 28.0 22.3 33.5









#35
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 23.149 min Scan# 1998
Delta R.T. -0.005 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 8670
Ion Ratio Lower Upper
252 100
253 24.0 19.4 29.0
125 10.7 8.2 12.4



#36
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 23.199 min Scan# 2012
Delta R.T. -0.005 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22

Tgt Ion:252 Resp: 9253
Ion Ratio Lower Upper
252 100
253 23.5 19.0 28.4
125 7.9 7.0 10.4









#37
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.809 min Scan# 2183
Delta R.T. -0.005 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	252	Resp:	8089
Ion Ratio	Lower	Upper	
252	100		
253	24.5	21.0	31.4
125	11.6	9.2	13.8



#38
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 26.595 min Scan# 2960
Delta R.T. -0.003 min
Lab File: BE098698.D
Acq: 1 Feb 2019 19:22


Tgt Ion:	278	Resp:	7655
Ion Ratio	Lower	Upper	
278	100		
139	15.3	12.7	19.1
279	26.1	22.0	33.0





#39
 Benzo(g,h,i)perylene
 Concen: 0.37 ng
 RT: 27.376 min Scan# 3179
 Delta R.T. -0.014 min
 Lab File: BE098698.D
 Acq: 1 Feb 2019 19:22

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:	276	Resp:	8212
Ion	Ratio	Lower	Upper
276	100		
277	25.5	20.3	30.5
138	16.9	13.3	19.9

