

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020419\Not Reviewed Data\
 Data File : BE098708.D
 Acq On : 4 Feb 2019 15:38
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Feb 05 01:26:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	947	0.40	ng	0.00
7) Naphthalene-d8	10.601	136	3897	0.40	ng	# 0.00
13) Acenaphthene-d10	14.472	164	2675	0.40	ng	0.00
19) Phenanthrene-d10	17.228	188	6272	0.40	ng	0.00
27) Chrysene-d12	21.415	240	6987	0.40	ng	0.00
34) Perylene-d12	23.919	264	6645	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.400	112	933	0.33	ng	0.00
5) Phenol-d6	7.001	99	1477	0.36	ng	0.00
8) Nitrobenzene-d5	8.977	82	1529	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	2961	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	445	0.38	ng	-0.01
15) 2-Fluorobiphenyl	13.089	172	4384	0.38	ng	-0.01
25) Fluoranthene-d10	19.259	212	34288	0.38	ng	0.00
29) Terphenyl-d14	19.857	244	6658	0.38	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.304	88	682	0.38	ng	93
3) n-Nitrosodimethylamine	3.661	42	623	0.30	ng	# 90
6) bis(2-Chloroethyl)ether	7.243	93	1021	0.36	ng	79
9) Naphthalene	10.653	128	17945	0.40	ng	100
10) Hexachlorobutadiene	10.939	225	1210	0.40	ng	95
12) 2-Methylnaphthalene	12.282	142	2862	0.39	ng	97
16) Acenaphthylene	14.199	152	5311	0.39	ng	99
17) Acenaphthene	14.530	154	3146	0.39	ng	99
18) Fluorene	15.514	166	4254	0.38	ng	99
21) Hexachlorobenzene	16.532	284	1742	0.40	ng	97
22) Pentachlorophenol	16.893	266	339	0.30	ng	97
23) Phenanthrene	17.268	178	6843	0.40	ng	99
24) Anthracene	17.361	178	5202	0.37	ng	98
26) Fluoranthene	19.288	202	8567	0.39	ng	99
28) Pyrene	19.650	202	8833	0.40	ng	99
30) Benzo(a)anthracene	21.391	228	8042	0.38	ng	99
31) Chrysene	21.451	228	8730	0.38	ng	99
32) Bis(2-ethylhexyl)phtha...	21.318	149	13146	0.34	ng	99
33) Indeno(1,2,3-cd)pyrene	26.566	276	9219	0.39	ng	99
35) Benzo(b)fluoranthene	23.149	252	8398	0.39	ng	100
36) Benzo(k)fluoranthene	23.199	252	8826	0.38	ng	98
37) Benzo(a)pyrene	23.812	252	7806	0.38	ng	99
38) Dibenzo(a,h)anthracene	26.591	278	7191	0.38	ng	99
39) Benzo(g,h,i)perylene	27.379	276	7889	0.38	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020419\Not Reviewed Data\
Data File : BE098708.D
Acq On : 4 Feb 2019 15:38
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Quant Time: Feb 05 01:26:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 24 14:44:44 2019
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.807 min Scan# 429
 Delta R.T. 0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38




Instrument :
 BNA_E
 ClientSampled :
 SSTDCCC0.4


Tgt Ion:152 Resp: 947
 Ion Ratio Lower Upper
 152 100
 150 148.4 121.4 182.2
 115 70.8 59.1 88.7



#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.304 min Scan# 38
 Delta R.T. -0.011 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38

Tgt Ion: 88 Resp: 682
 Ion Ratio Lower Upper
 88 100
 58 85.8 73.9 110.9
 43 49.1 43.7 65.5









#3
n-Nitrosodimethylamine
Concen: 0.30 ng
RT: 3.661 min Scan# 69
Delta R.T. 0.012 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 42 Resp: 623




Ion	Ratio	Lower	Upper
42	100		
74	81.7	58.9	88.3
44	0.0	0.0	0.0




#4
2-Fluorophenol
Concen: 0.33 ng
RT: 5.400 min Scan# 220
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion: 112 Resp: 933

Ion	Ratio	Lower	Upper
112	100		
64	86.2	62.9	94.3
63	178.8	152.2	228.4









#5
Phenol-d6
Concen: 0.36 ng
RT: 7.001 min Scan# 359
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 1477




Ion	Ratio	Lower	Upper
99	100		
42	33.4	24.6	37.0
71	44.3	35.1	52.7




#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.243 min Scan# 380
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion: 93 Resp: 1021

Ion	Ratio	Lower	Upper
93	100		
63	286.4	267.3	400.9
95	30.4	26.2	39.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.601 min Scan# 649
Delta R.T. -0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	3897
Ion	Ratio	Lower	Upper
136	100		
137	12.3	9.7	14.5
54	54.6	35.5	53.3#
68	8.5	5.0	7.6#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.977 min Scan# 524
Delta R.T. -0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:	82	Resp:	1529
Ion	Ratio	Lower	Upper
82	100		
128	141.4	96.5	144.7
54	293.3	224.7	337.1









#9
Naphthalene
Concen: 0.40 ng
RT: 10.653 min Scan# 653
Delta R.T. -0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128	Resp:	17945
Ion Ratio	Lower	Upper
128	100	
129	3.2	2.6 3.8
127	3.9	3.0 4.6



#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.939 min Scan# 675
Delta R.T. -0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:225	Resp:	1210
Ion Ratio	Lower	Upper
225	100	
223	64.3	47.8 71.8
227	66.1	50.5 75.7









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.210 min Scan# 763
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 2961
Ion Ratio Lower Upper
152 100
151 19.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.282 min Scan# 768
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38


Tgt Ion:142 Resp: 2862
Ion Ratio Lower Upper
142 100
141 88.7 70.0 105.0
115 48.3 35.2 52.8








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.472 min Scan# 920
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4





Tgt Ion:164 Resp: 2675
Ion Ratio Lower Upper
164 100
162 97.4 80.5 120.7
160 48.5 40.1 60.1




#14
2,4,6-Tribromophenol
Concen: 0.38 ng
RT: 15.969 min Scan# 1026
Delta R.T. -0.013 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38



Tgt Ion:330 Resp: 445
Ion Ratio Lower Upper
330 100
332 98.4 73.3 109.9
141 41.8 33.1 49.7









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 13.089 min Scan# 824
Delta R.T. -0.014 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	4384
Ion Ratio	Lower	Upper
172	100	
171	36.0	29.0 43.4
170	25.7	20.8 31.2



#16
Acenaphthylene
Concen: 0.39 ng
RT: 14.199 min Scan# 901
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:152	Resp:	5311
Ion Ratio	Lower	Upper
152	100	
151	19.8	16.3 24.5
153	12.9	10.7 16.1









#17
Acenaphthene
Concen: 0.39 ng
RT: 14.530 min Scan# 924
Delta R.T. -0.013 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 3146
Ion Ratio Lower Upper
154 100
153 116.1 93.9 140.9
152 57.8 45.8 68.8



#18
Fluorene
Concen: 0.38 ng
RT: 15.514 min Scan# 992
Delta R.T. 0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:166 Resp: 4254
Ion Ratio Lower Upper
166 100
165 100.5 79.3 118.9
167 13.0 10.6 15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.228 min Scan# 1120
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 6272
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 4.7 7.1



#21
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.532 min Scan# 1068
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:284 Resp: 1742
Ion Ratio Lower Upper
284 100
142 33.7 25.5 38.3
249 35.0 26.6 39.8









#22
 Pentachlorophenol
 Concen: 0.30 ng
 RT: 16.893 min Scan# 1095
 Delta R.T. -0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	266	Resp:	339
Ion	Ratio	Lower	Upper
266	100		
264	69.0	53.7	80.5
268	78.0	59.8	89.8



#23
 Phenanthrene
 Concen: 0.40 ng
 RT: 17.268 min Scan# 1123
 Delta R.T. -0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38

Tgt Ion:	178	Resp:	6843
Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.9	23.9
179	14.6	12.2	18.4









#24
 Anthracene
 Concen: 0.37 ng
 RT: 17.361 min Scan# 1130
 Delta R.T. 0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:178 Resp: 5202
 Ion Ratio Lower Upper
 178 100
 176 19.2 14.9 22.3
 179 15.9 11.8 17.8



#25
 Fluoranthene-d10
 Concen: 0.38 ng
 RT: 19.259 min Scan# 1391
 Delta R.T. 0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38

Tgt Ion:212 Resp: 34288
 Ion Ratio Lower Upper
 212 100
 106 2.2 1.7 2.5
 104 1.2 1.0 1.6









#26
Fluoranthene
Concen: 0.39 ng
RT: 19.288 min Scan# 1396
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:202 Resp: 8567
Ion Ratio Lower Upper
202 100
101 8.0 6.7 10.1
203 17.0 13.3 19.9



#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.415 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:240 Resp: 6987
Ion Ratio Lower Upper
240 100
120 6.1 4.7 7.1
236 26.6 21.9 32.9









#28
Pyrene
Concen: 0.40 ng
RT: 19.650 min Scan# 1459
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:202	Resp:	8833
Ion Ratio	Lower	Upper
202	100	
200	21.5	17.0 25.6
203	16.8	13.8 20.8



#29
Terphenyl-d14
Concen: 0.38 ng
RT: 19.857 min Scan# 1495
Delta R.T. 0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:244	Resp:	6658
Ion Ratio	Lower	Upper
244	100	
212	37.1	29.4 44.2
122	8.5	7.1 10.7









#30
Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.391 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228	Resp:	8042	
Ion	Ratio	Lower	Upper
228	100		
226	28.5	23.0	34.4
229	19.9	15.0	22.6



#31
Chrysene
Concen: 0.38 ng
RT: 21.451 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:228	Resp:	8730	
Ion	Ratio	Lower	Upper
228	100		
226	31.3	24.6	36.8
229	20.4	16.4	24.6









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.34 ng
 RT: 21.318 min Scan# 1626
 Delta R.T. -0.000 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:149	Resp:	13146
Ion Ratio	Lower	Upper
149	100	
167	7.1	5.4 8.2
279	1.2	1.0 1.4



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.39 ng
 RT: 26.566 min Scan# 2952
 Delta R.T. -0.010 min
 Lab File: BE098708.D
 Acq: 4 Feb 2019 15:38

Tgt Ion:276	Resp:	9219
Ion Ratio	Lower	Upper
276	100	
138	16.2	13.3 19.9
277	24.5	19.4 29.0





#34

Perylene-d12

Concen: 0.40 ng

RT: 23.919 min Scan# 2214

Delta R.T. -0.008 min

Lab File: BE098708.D

Acq: 4 Feb 2019 15:38





Tgt Ion:264 Resp: 6645

Ion Ratio Lower Upper

264 100

260 26.0 20.5 30.7

265 28.6 22.3 33.5



#35

Benzo(b)fluoranthene

Concen: 0.39 ng

RT: 23.149 min Scan# 1998

Delta R.T. -0.004 min

Lab File: BE098708.D

Acq: 4 Feb 2019 15:38




Tgt Ion:252 Resp: 8398


Ion Ratio Lower Upper

252 100

253 24.2 19.4 29.0

125 9.8 8.2 12.4









#36
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 23.199 min Scan# 2012
Delta R.T. -0.004 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 8826
Ion Ratio Lower Upper
252 100
253 24.8 19.0 28.4
125 8.2 7.0 10.4



#37
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.812 min Scan# 2184
Delta R.T. -0.001 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:252 Resp: 7806
Ion Ratio Lower Upper
252 100
253 25.5 21.0 31.4
125 11.6 9.2 13.8









#38
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 26.591 min Scan# 2959
Delta R.T. -0.007 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:278 Resp: 7191
Ion Ratio Lower Upper
278 100
139 15.6 12.7 19.1
279 26.7 22.0 33.0



#39
Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 27.379 min Scan# 3180
Delta R.T. -0.010 min
Lab File: BE098708.D
Acq: 4 Feb 2019 15:38

Tgt Ion:276 Resp: 7889
Ion Ratio Lower Upper
276 100
277 26.0 20.3 30.5
138 17.1 13.3 19.9

