

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020419\
 Data File : BE098706.D
 Acq On : 4 Feb 2019 13:44
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 04 15:24:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.739	0.655	11.4	102	0.00
3	n-Nitrosodimethylamine	0.877	0.805	8.2	97	0.00
4 S	2-Fluorophenol	1.182	1.000	15.4	91	0.00
5 S	Phenol-d6	1.737	1.549	10.8	102	0.00
6	bis(2-Chloroethyl)ether	1.183	0.905	23.5	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.389	0.376	3.3	105	0.00
9	Naphthalene	4.584	4.541	0.9	111	0.00
10	Hexachlorobutadiene	0.307	0.294	4.2	106	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.750	2.0	108	0.00
12	2-Methylnaphthalene	0.761	0.725	4.7	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.177	0.187	-5.6	116	-0.01
15 S	2-Fluorobiphenyl	1.714	1.655	3.4	110	0.00
16	Acenaphthylene	2.052	1.991	3.0	111	0.00
17	Acenaphthene	1.212	1.169	3.5	106	-0.02
18	Fluorene	1.656	1.610	2.8	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.236	0.230	2.5	112	-0.01
21	Hexachlorobenzene	0.278	0.274	1.4	111	0.00
22	Pentachlorophenol	0.071	0.062	12.7	108	0.00
23	Phenanthrene	1.101	1.067	3.1	111	0.00
24	Anthracene	0.919	0.848	7.7	108	0.00
25 SURR	Fluoranthene-d10	5.713	5.475	4.2	109	0.00
26	Fluoranthene	1.417	1.374	3.0	111	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	-0.01
28	Pyrene	1.274	1.265	0.7	110	0.00
29 S	Terphenyl-d14	0.998	0.964	3.4	106	0.00
30	Benzo(a)anthracene	1.207	1.110	8.0	106	0.00
31	Chrysene	1.308	1.327	-1.5	113	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	2.086	6.8	106	-0.01
33	Indeno(1,2,3-cd)pyrene	1.370	1.350	1.5	116	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	113	0.00
35	Benzo(b)fluoranthene	1.294	1.232	4.8	109	0.00
36	Benzo(k)fluoranthene	1.383	1.284	7.2	108	0.00
37 C	Benzo(a)pyrene	1.237	1.216	1.7	115	0.00
38	Dibenzo(a,h)anthracene	1.146	1.072	6.5	115	0.00
39	Benzo(g,h,i)perylene	1.258	1.175	6.6	110	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0