

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020419\
 Data File : BE098715.D
 Acq On : 4 Feb 2019 19:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 04 23:59:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	0.00
2	1,4-Dioxane	0.400	0.396	1.0	126	-0.01
3	n-Nitrosodimethylamine	0.400	0.408	-2.0	127	0.00
4 S	2-Fluorophenol	0.400	0.392	2.0	114	0.00
5 S	Phenol-d6	0.400	0.386	3.5	120	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	125	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.385	3.8	114	0.00
9	Naphthalene	0.400	0.402	-0.5	122	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	118	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	116	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.400	0.453	-13.2	136	-0.01
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	122	-0.01
16	Acenaphthylene	0.400	0.386	3.5	121	0.00
17	Acenaphthene	0.400	0.393	1.8	118	-0.01
18	Fluorene	0.400	0.399	0.3	125	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	122	0.00
20	4-Bromophenyl-phenylether	0.400	0.378	5.5	121	-0.01
21	Hexachlorobenzene	0.400	0.391	2.3	122	0.00
22	Pentachlorophenol	0.400	0.468	-17.0	159	0.00
23	Phenanthrene	0.400	0.402	-0.5	128	0.00
24	Anthracene	0.400	0.389	2.8	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.370	7.5	117	0.00
26	Fluoranthene	0.400	0.398	0.5	126	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	120	-0.01
28	Pyrene	0.400	0.407	-1.7	126	0.00
29 S	Terphenyl-d14	0.400	0.424	-6.0	130	0.00
30	Benzo(a)anthracene	0.400	0.425	-6.2	137	0.00
31	Chrysene	0.400	0.408	-2.0	127	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.410	-2.5	130	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.417	-4.2	137	0.00
34 I	Perylene-d12	0.400	0.400	0.0	125	0.00
35	Benzo(b)fluoranthene	0.400	0.403	-0.8	128	0.00
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	134	0.00
37 C	Benzo(a)pyrene	0.400	0.405	-1.3	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.411	-2.7	141	-0.01
39	Benzo(g,h,i)perylene	0.400	0.388	3.0	127	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0