

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020419\
 Data File : BE098715.D
 Acq On : 4 Feb 2019 19:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 04 23:59:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.739	0.740	-0.1	126	-0.01
3	n-Nitrosodimethylamine	0.877	0.969	-10.5	127	0.00
4 S	2-Fluorophenol	1.182	1.158	2.0	114	0.00
5 S	Phenol-d6	1.737	1.679	3.3	120	-0.01
6	bis(2-Chloroethyl)ether	1.183	1.144	3.3	125	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.389	0.374	3.9	114	0.00
9	Naphthalene	4.584	4.604	-0.4	122	0.00
10	Hexachlorobutadiene	0.307	0.300	2.3	118	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.744	2.7	116	0.00
12	2-Methylnaphthalene	0.761	0.748	1.7	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.177	0.200	-13.0	136	-0.01
15 S	2-Fluorobiphenyl	1.714	1.684	1.8	122	-0.01
16	Acenaphthylene	2.052	1.981	3.5	121	0.00
17	Acenaphthene	1.212	1.191	1.7	118	-0.01
18	Fluorene	1.656	1.652	0.2	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
20	4-Bromophenyl-phenylether	0.236	0.223	5.5	121	-0.01
21	Hexachlorobenzene	0.278	0.272	2.2	122	0.00
22	Pentachlorophenol	0.071	0.083	-16.9	159#	0.00
23	Phenanthrene	1.101	1.106	-0.5	128	0.00
24	Anthracene	0.919	0.892	2.9	126	0.00
25 SURR	Fluoranthene-d10	5.713	5.287	7.5	117	0.00
26	Fluoranthene	1.417	1.410	0.5	126	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	120	-0.01
28	Pyrene	1.274	1.297	-1.8	126	0.00
29 S	Terphenyl-d14	0.998	1.059	-6.1	130	0.00
30	Benzo(a)anthracene	1.207	1.282	-6.2	137	0.00
31	Chrysene	1.308	1.333	-1.9	127	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	2.297	-2.6	130	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.430	-4.4	137	0.00
34 I	Perylene-d12	1.000	1.000	0.0	125	0.00
35	Benzo(b)fluoranthene	1.294	1.305	-0.9	128	0.00
36	Benzo(k)fluoranthene	1.383	1.428	-3.3	134	0.00
37 C	Benzo(a)pyrene	1.237	1.253	-1.3	132	0.00
38	Dibenzo(a,h)anthracene	1.146	1.177	-2.7	141	-0.01
39	Benzo(g,h,i)perylene	1.258	1.222	2.9	127	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0