

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020719\
 Data File : BE098717.D
 Acq On : 7 Feb 2019 9:19
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 07 10:31:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	0.01
2	1,4-Dioxane	0.400	0.345	13.8	110	0.00
3	n-Nitrosodimethylamine	0.400	0.353	11.8	104	0.01
4 S	2-Fluorophenol	0.400	0.365	8.8	104	0.01
5 S	Phenol-d6	0.400	0.355	11.3	107	0.01
6	bis(2-Chloroethyl)ether	0.400	0.341	14.8	107	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.01
8 S	Nitrobenzene-d5	0.400	0.376	6.0	107	0.01
9	Naphthalene	0.400	0.398	0.5	118	0.01
10	Hexachlorobutadiene	0.400	0.399	0.3	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	113	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	114	0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.02
14 S	2,4,6-Tribromophenol	0.400	0.348	13.0	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.374	6.5	114	0.00
16	Acenaphthylene	0.400	0.381	4.8	117	0.00
17	Acenaphthene	0.400	0.391	2.3	116	0.00
18	Fluorene	0.400	0.379	5.3	116	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.400	0.375	6.3	112	0.00
21	Hexachlorobenzene	0.400	0.390	2.5	114	0.00
22	Pentachlorophenol	0.400	0.497	-24.2	158	0.00
23	Phenanthrene	0.400	0.392	2.0	116	0.00
24	Anthracene	0.400	0.371	7.3	111	0.00
25 SURR	Fluoranthene-d10	0.400	0.385	3.8	114	0.00
26	Fluoranthene	0.400	0.389	2.8	116	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
28	Pyrene	0.400	0.399	0.3	115	0.00
29 S	Terphenyl-d14	0.400	0.383	4.3	109	0.00
30	Benzo(a)anthracene	0.400	0.365	8.8	110	0.00
31	Chrysene	0.400	0.384	4.0	111	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.343	14.2	101	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.349	12.8	107	0.00
34 I	Perylene-d12	0.400	0.400	0.0	116	0.00
35	Benzo(b)fluoranthene	0.400	0.383	4.3	112	0.00
36	Benzo(k)fluoranthene	0.400	0.387	3.3	116	0.00
37 C	Benzo(a)pyrene	0.400	0.381	4.8	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.327	18.3	104	0.00
39	Benzo(g,h,i)perylene	0.400	0.356	11.0	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0