

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020719\
 Data File : BE098725.D
 Acq On : 7 Feb 2019 14:38
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 09 06:14:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.01
2	1,4-Dioxane	0.400	0.409	-2.2	121	-0.01
3	n-Nitrosodimethylamine	0.400	0.418	-4.5	123	0.00
4 S	2-Fluorophenol	0.400	0.411	-2.7	113	0.01
5 S	Phenol-d6	0.400	0.391	2.3	115	0.01
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	128	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.01
8 S	Nitrobenzene-d5	0.400	0.375	6.3	110	0.01
9	Naphthalene	0.400	0.390	2.5	118	0.01
10	Hexachlorobutadiene	0.400	0.388	3.0	116	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	112	0.01
12	2-Methylnaphthalene	0.400	0.377	5.8	113	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.01
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	116	0.00
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	116	0.00
16	Acenaphthylene	0.400	0.379	5.3	116	0.00
17	Acenaphthene	0.400	0.388	3.0	114	0.00
18	Fluorene	0.400	0.383	4.3	117	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	114	0.00
21	Hexachlorobenzene	0.400	0.375	6.3	111	0.00
22	Pentachlorophenol	0.400	0.369	7.8	120	0.00
23	Phenanthrene	0.400	0.396	1.0	120	0.00
24	Anthracene	0.400	0.387	3.3	119	0.00
25 SURR	Fluoranthene-d10	0.400	0.384	4.0	115	0.00
26	Fluoranthene	0.400	0.385	3.8	116	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
28	Pyrene	0.400	0.399	0.3	118	0.00
29 S	Terphenyl-d14	0.400	0.388	3.0	113	0.00
30	Benzo(a)anthracene	0.400	0.376	6.0	116	0.01
31	Chrysene	0.400	0.392	2.0	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.360	10.0	108	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	117	0.00
34 I	Perylene-d12	0.400	0.400	0.0	117	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	115	0.00
36	Benzo(k)fluoranthene	0.400	0.391	2.3	118	0.00
37 C	Benzo(a)pyrene	0.400	0.379	5.3	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.361	9.8	115	0.00
39	Benzo(g,h,i)perylene	0.400	0.370	7.5	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0