

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE021319\
 Data File : BE098727.D
 Acq On : 13 Feb 2019 14:56
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 14 04:05:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	119	0.00
2	1,4-Dioxane	0.400	0.419	-4.7	132	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	106	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	107	0.00
5 S	Phenol-d6	0.400	0.361	9.8	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	132	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	120	0.00
8 S	Nitrobenzene-d5	0.400	0.392	2.0	118	0.00
9	Naphthalene	0.400	0.393	1.8	121	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	115	0.00
12	2-Methylnaphthalene	0.400	0.370	7.5	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.400	0.316	21.0	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	120	0.00
16	Acenaphthylene	0.400	0.373	6.8	114	0.00
17	Acenaphthene	0.400	0.393	1.8	116	0.00
18	Fluorene	0.400	0.374	6.5	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.400	0.363	9.3	108	0.00
21	Hexachlorobenzene	0.400	0.392	2.0	114	0.00
22	Pentachlorophenol	0.400	0.443	-10.7	141	0.00
23	Phenanthrene	0.400	0.391	2.3	116	0.00
24	Anthracene	0.400	0.366	8.5	109	0.00
25 SURR	Fluoranthene-d10	0.400	0.394	1.5	117	0.00
26	Fluoranthene	0.400	0.398	0.5	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
28	Pyrene	0.400	0.415	-3.7	120	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	111	0.00
30	Benzo(a)anthracene	0.400	0.381	4.8	115	0.01
31	Chrysene	0.400	0.389	2.8	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.387	3.3	114	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.309	22.8	94	0.00
34 I	Perylene-d12	0.400	0.400	0.0	116	0.00
35	Benzo(b)fluoranthene	0.400	0.381	4.8	112	0.00
36	Benzo(k)fluoranthene	0.400	0.390	2.5	117	0.00
37 C	Benzo(a)pyrene	0.400	0.370	7.5	112	0.00
38	Dibenzo(a,h)anthracene	0.400	0.258	35.5#	82	0.00
39	Benzo(g,h,i)perylene	0.400	0.329	17.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0