

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE021319\
 Data File : BE098733.D
 Acq On : 13 Feb 2019 18:40
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 14 04:07:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.739	0.690	6.6	116	0.00
3	n-Nitrosodimethylamine	0.877	0.715	18.5	93	0.00
4 S	2-Fluorophenol	1.182	1.009	14.6	99	0.00
5 S	Phenol-d6	1.737	1.504	13.4	107	0.00
6	bis(2-Chloroethyl)ether	1.183	1.024	13.4	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.389	0.356	8.5	107	0.00
9	Naphthalene	4.584	4.569	0.3	120	0.00
10	Hexachlorobutadiene	0.307	0.297	3.3	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.730	4.6	112	0.00
12	2-Methylnaphthalene	0.761	0.709	6.8	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.177	0.167	5.6	110	-0.01
15 S	2-Fluorobiphenyl	1.714	1.673	2.4	118	-0.01
16	Acenaphthylene	2.052	1.920	6.4	114	0.00
17	Acenaphthene	1.212	1.186	2.1	115	-0.01
18	Fluorene	1.656	1.617	2.4	119	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.236	0.216	8.5	111	-0.01
21	Hexachlorobenzene	0.278	0.271	2.5	116	0.00
22	Pentachlorophenol	0.071	0.050	29.6#	91	0.00
23	Phenanthrene	1.101	1.065	3.3	117	0.00
24	Anthracene	0.919	0.847	7.8	114	0.00
25 SURR	Fluoranthene-d10	5.713	5.463	4.4	115	0.00
26	Fluoranthene	1.417	1.370	3.3	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
28	Pyrene	1.274	1.236	3.0	118	0.00
29 S	Terphenyl-d14	0.998	0.946	5.2	114	0.00
30	Benzo(a)anthracene	1.207	1.124	6.9	118	0.00
31	Chrysene	1.308	1.242	5.0	116	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	1.916	14.4	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.198	12.6	112	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.294	1.219	5.8	114	0.00
36	Benzo(k)fluoranthene	1.383	1.309	5.4	118	0.00
37 C	Benzo(a)pyrene	1.237	1.158	6.4	117	0.00
38	Dibenzo(a,h)anthracene	1.146	0.928	19.0	107	0.00
39	Benzo(g,h,i)perylene	1.258	1.109	11.8	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0