

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022019\
 Data File : BE098735.D
 Acq On : 20 Feb 2019 10:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 14:03:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	0.00
2	1,4-Dioxane	0.400	0.363	9.3	115	0.00
3	n-Nitrosodimethylamine	0.400	0.414	-3.5	126	0.00
4 S	2-Fluorophenol	0.400	0.333	16.8	95	0.01
5 S	Phenol-d6	0.400	0.330	17.5	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	109	0.00
9	Naphthalene	0.400	0.393	1.8	116	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	110	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.400	0.321	19.8	93	-0.01
15 S	2-Fluorobiphenyl	0.400	0.379	5.3	114	-0.01
16	Acenaphthylene	0.400	0.369	7.8	112	-0.01
17	Acenaphthene	0.400	0.381	4.8	111	-0.01
18	Fluorene	0.400	0.380	5.0	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	-0.01
20	4-Bromophenyl-phenylether	0.400	0.373	6.8	114	-0.01
21	Hexachlorobenzene	0.400	0.382	4.5	114	0.00
22	Pentachlorophenol	0.400	0.363	9.3	118	0.00
23	Phenanthrene	0.400	0.388	3.0	118	0.00
24	Anthracene	0.400	0.377	5.8	116	0.00
25 SURR	Fluoranthene-d10	0.400	0.396	1.0	120	0.00
26	Fluoranthene	0.400	0.400	0.0	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	113	-0.01
28	Pyrene	0.400	0.424	-6.0	123	0.00
29 S	Terphenyl-d14	0.400	0.397	0.8	114	0.00
30	Benzo(a)anthracene	0.400	0.409	-2.2	124	0.00
31	Chrysene	0.400	0.386	3.5	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.407	-1.7	121	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.268	33.0#	83	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	118	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	116	0.00
36	Benzo(k)fluoranthene	0.400	0.388	3.0	119	0.00
37 C	Benzo(a)pyrene	0.400	0.379	5.3	116	0.00
38	Dibenzo(a,h)anthracene	0.400	0.224	44.0#	72	0.00
39	Benzo(g,h,i)perylene	0.400	0.321	19.8	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0