


Nitrobenzene-d5

Response Ratio




$$\text{Response} = 4.834\text{e-}001 * \text{Amt} - 8.574\text{e-}002$$

Coef of Det (r^2) = 0.995466 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M


Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

2,4,6-Tribromophenol



Response = 2.165e-001 * Amt - 4.934e-002
 Coef of Det (r^2) = 0.992099 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M
 Calibration Table Last Updated: Tue Feb 26 15:44:42 2019


Pentachlorophenol



Response = $8.749 \times 10^{-2} \times \text{Amt} - 4.591 \times 10^{-2}$
 Coef of Det (r^2) = 0.991419 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M
 Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

Anthracene

Response Ratio



$$\text{Response} = 1.114\text{e}+000 * \text{Amt} - 2.319\text{e}-001$$


Coef of Det (r^2) = 0.998696 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M

Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

Indeno(1,2,3-cd)pyrene

Response Ratio



$$\text{Response} = 1.342\text{e}+000 * \text{Amt} - 3.024\text{e}-001$$


Coef of Det (r^2) = 0.999007 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M

Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

Dibenzo (a,h) anthracene

Response Ratio



$$\text{Response} = 1.128\text{e}+000 * \text{Amt} - 4.073\text{e}-001$$


Coef of Det (r^2) = 0.997012 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M

Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

n-Nitrosodimethylamine

Response Ratio



$$\text{Response} = 1.024\text{e}+000 * \text{Amt} - 2.385\text{e}-001$$


Coef of Det (r^2) = 0.999356 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M

Calibration Table Last Updated: Tue Feb 26 15:44:42 2019

bis(2-Chloroethyl) ether

Response Ratio



$$\text{Response} = 1.309\text{e}+000 * \text{Amt} - 2.247\text{e}-001$$

Coef of Det (r^2) = 0.998476 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE022619.M

Calibration Table Last Updated: Tue Feb 26 15:44:42 2019