

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022619\
 Data File : BE098769.D
 Acq On : 26 Feb 2019 21:17
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 00:24:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 26 15:44:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	178#	-0.01
2	1,4-Dioxane	0.690	0.528	23.5	137	0.00
3	n-Nitrosodimethylamine	0.758	0.948	-25.1#	219#	-0.02
4 S	2-Fluorophenol	1.043	0.907	13.0	158#	-0.01
5 S	Phenol-d6	1.560	1.896	-21.5	222#	-0.01
6	bis(2-Chloroethyl)ether	1.071	1.748	-63.2#	295#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	275#	0.00
8 S	Nitrobenzene-d5	0.380	0.408	-7.4	297#	-0.01
9	Naphthalene	4.611	4.455	3.4	274#	-0.01
10	Hexachlorobutadiene	0.306	0.284	7.2	262#	0.00
11 SURR	2-Methylnaphthalene-d10	0.757	0.803	-6.1	306#	-0.01
12	2-Methylnaphthalene	0.735	0.809	-10.1	321#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	313#	0.00
14 S	2,4,6-Tribromophenol	0.158	0.199	-25.9#	391#	-0.01
15 S	2-Fluorobiphenyl	1.693	1.660	1.9	308#	0.00
16	Acenaphthylene	2.045	2.034	0.5	324#	-0.01
17	Acenaphthene	1.209	1.180	2.4	311#	0.00
18	Fluorene	1.673	1.669	0.2	324#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	304#	-0.01
20	4-Bromophenyl-phenylether	0.225	0.238	-5.8	348#	0.00
21	Hexachlorobenzene	0.270	0.265	1.9	298#	-0.01
22	Pentachlorophenol	0.061	0.065	-6.6	398#	-0.01
23	Phenanthrene	1.096	1.103	-0.6	319#	-0.01
24	Anthracene	0.885	0.957	-8.1	353#	-0.01
25 SURR	Fluoranthene-d10	5.692	5.512	3.2	304#	0.00
26	Fluoranthene	1.414	1.383	2.2	310#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	371#	-0.01
28	Pyrene	1.329	1.073	19.3	316#	0.00
29 S	Terphenyl-d14	0.992	0.836	15.7	327#	0.00
30	Benzo(a)anthracene	1.152	0.938	18.6	331#	0.00
31	Chrysene	1.300	1.011	22.2	287#	-0.01
32	Bis(2-ethylhexyl)phthalate	2.456	2.257	8.1	376#	0.00
33	Indeno(1,2,3-cd)pyrene	1.032	0.869	15.8	305#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	299#	0.00
35	Benzo(b)fluoranthene	1.260	1.177	6.6	284#	0.00
36	Benzo(k)fluoranthene	1.313	1.185	9.7	276#	0.00
37 C	Benzo(a)pyrene	1.163	1.072	7.8	283#	0.00
38	Dibenzo(a,h)anthracene	0.739	0.873	-18.1	374#	-0.02
39	Benzo(g,h,i)perylene	1.085	0.952	12.3	266#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0