

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022619\
 Data File : BE098769.D
 Acq On : 26 Feb 2019 21:17
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 00:24:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 26 15:44:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	178	-0.01
2	1,4-Dioxane	0.400	0.306	23.5	137	0.00
3	n-Nitrosodimethylamine	0.400	0.463	-15.8	219	-0.02
4 S	2-Fluorophenol	0.400	0.348	13.0	158	-0.01
5 S	Phenol-d6	0.400	0.486	-21.5	222	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.603	-50.7#	295	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	275	0.00
8 S	Nitrobenzene-d5	0.400	0.408	-2.0	297	-0.01
9	Naphthalene	0.400	0.386	3.5	274	-0.01
10	Hexachlorobutadiene	0.400	0.371	7.3	262	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.424	-6.0	306	-0.01
12	2-Methylnaphthalene	0.400	0.440	-10.0	321	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	313	0.00
14 S	2,4,6-Tribromophenol	0.400	0.460	-15.0	391	-0.01
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	308	0.00
16	Acenaphthylene	0.400	0.398	0.5	324	-0.01
17	Acenaphthene	0.400	0.391	2.3	311	0.00
18	Fluorene	0.400	0.399	0.3	324	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	304	-0.01
20	4-Bromophenyl-phenylether	0.400	0.424	-6.0	348	0.00
21	Hexachlorobenzene	0.400	0.392	2.0	298	-0.01
22	Pentachlorophenol	0.400	0.507	-26.7#	398	-0.01
23	Phenanthrene	0.400	0.403	-0.8	319	-0.01
24	Anthracene	0.400	0.427	-6.7	353	-0.01
25 SURR	Fluoranthene-d10	0.400	0.387	3.3	304	0.00
26	Fluoranthene	0.400	0.391	2.3	310	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	371	-0.01
28	Pyrene	0.400	0.323	19.3	316	0.00
29 S	Terphenyl-d14	0.400	0.337	15.8	327	0.00
30	Benzo(a)anthracene	0.400	0.326	18.5	331	0.00
31	Chrysene	0.400	0.311	22.3	287	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.368	8.0	376	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.349	12.8	305	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	299	0.00
35	Benzo(b)fluoranthene	0.400	0.373	6.8	284	0.00
36	Benzo(k)fluoranthene	0.400	0.361	9.8	276	0.00
37 C	Benzo(a)pyrene	0.400	0.369	7.8	283	0.00
38	Dibenzo(a,h)anthracene	0.400	0.454	-13.5	374	-0.02
39	Benzo(g,h,i)perylene	0.400	0.351	12.3	266	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0