

Data Path : \\TERASTORAGE\Terastorage\svoasrv\HPCHEM1\BNA_E\Data\BE022619\
 Data File : BE098757.D
 Acq On : 26 Feb 2019 12:41
 Operator : JU/SJ
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

Sohil
 2/27/2019 4:22:42 PM

Quant Time: Feb 26 13:33:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 26 13:31:15 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.815	152	944	0.40	ng	0.00
7) Naphthalene-d8	10.602	136	4144	0.40	ng	0.00
13) Acenaphthene-d10	14.472	164	2586	0.40	ng	0.00
19) Phenanthrene-d10	17.227	188	6289	0.40	ng	# 0.00
27) Chrysene-d12	21.415	240	6892	0.40	ng	0.00
34) Perylene-d12	23.921	264	6594	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.408	112	967	0.38	ng	0.00
5) Phenol-d6	6.998	99	1439	0.40	ng	0.00
8) Nitrobenzene-d5	8.978	82	1564	0.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	2988	0.44	ng	0.00
14) 2,4,6-Tribromophenol	15.982	330	413	0.37	ng	0.00
15) 2-Fluorobiphenyl	13.089	172	4368	0.41	ng	0.00
25) Fluoranthene-d10	19.253	212	34694	0.45	ng	0.00
29) Terphenyl-d14	19.857	244	6537	0.43	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.312	88	648	0.35	ng	97
3) n-Nitrosodimethylamine	3.669	42	730	0.53	ng	# 82
6) bis(2-Chloroethyl)ether	7.240	93	1000	0.37	ng	93
9) Naphthalene	10.654	128	18533	0.45	ng	100
10) Hexachlorobutadiene	10.927	225	1237	0.42	ng	97
12) 2-Methylnaphthalene	12.282	142	2871	0.42	ng	94
16) Acenaphthylene	14.198	152	5079	0.45	ng	98
17) Acenaphthene	14.529	154	3072	0.44	ng	99
18) Fluorene	15.514	166	4175	0.46	ng	97
20) 4-Bromophenyl-phenylether	16.411	248	1313	0.39	ng	# 84
21) Hexachlorobenzene	16.531	284	1700	0.42	ng	95
22) Pentachlorophenol	16.893	266	312	0.27	ng	# 89
23) Phenanthrene	17.268	178	6633	0.45	ng	100
24) Anthracene	17.361	178	5189	0.41	ng	97
26) Fluoranthene	19.288	202	8536	0.45	ng	99
28) Pyrene	19.650	202	8683	0.46	ng	99
30) Benzo(a)anthracene	21.391	228	7261	0.39	ng	97
31) Chrysene	21.451	228	9014	0.47	ng	98
32) Bis(2-ethylhexyl)phtha...	21.306	149	15353	0.39	ng	99
33) Indeno(1,2,3-cd)pyrene	26.570	276	7298	0.37	ng	99
35) Benzo(b)fluoranthene	23.147	252	8173m	0.45	ng	
36) Benzo(k)fluoranthene	23.197	252	8461	0.43	ng	98
37) Benzo(a)pyrene	23.807	252	7455	0.42	ng	# 97
38) Dibenzo(a,h)anthracene	26.595	278	4597	0.29	ng	96
39) Benzo(g,h,i)perylene	27.373	276	7063	0.41	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : \\TERASTORAGE\Terastorage\svoasrv\HPCHEM1\BNA_E\Data\BE022619\
Data File : BE098757.D
Acq On : 26 Feb 2019 12:41
Operator : JU/SJ
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED

Sohil
2/27/2019 4:22:42 PM


Quant Time: Feb 26 13:33:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 26 13:31:15 2019
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.815 min Scan# 430
 Delta R.T. 0.000 min
 Lab File: BE098757.D
 Acq: 26 Feb 2019 12:41




Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.4




Tgt Ion: 152 Resp: 944
 Ion Ratio Lower Upper
 152 100
 150 142.9 121.4 182.2
 115 67.6 59.1 88.7

Manual Integrations
 APPROVED



Sohil
 2/27/2019 4:22:42 PM




#2
 1,4-Dioxane
 Concen: 0.35 ng
 RT: 3.312 min Scan# 39
 Delta R.T. 0.000 min
 Lab File: BE098757.D
 Acq: 26 Feb 2019 12:41



Tgt Ion: 88 Resp: 648
 Ion Ratio Lower Upper
 88 100
 58 96.5 73.9 110.9
 43 53.1 43.7 65.5









#3
n-Nitrosodimethylamine
Concen: 0.53 ng
RT: 3.669 min Scan# 70
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion: 42 Resp: 730
Ion Ratio Lower Upper
42 100
74 58.8 58.9 88.3#
44 0.0 0.0 0.0

Instrument :
BNA_E
ClientSampled :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#4
2-Fluorophenol
Concen: 0.38 ng
RT: 5.408 min Scan# 221
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion: 112 Resp: 967
Ion Ratio Lower Upper
112 100
64 86.0 62.9 94.3
63 194.1 152.2 228.4









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.998 min Scan# 359
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	99	Resp:	1439
Ion	Ratio	Lower	Upper
99	100		
42	29.8	24.6	37.0
71	49.1	35.1	52.7

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.37 ng
RT: 7.240 min Scan# 380
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41


Tgt Ion:	93	Resp:	1000
Ion	Ratio	Lower	Upper
93	100		
63	318.3	267.3	400.9
95	31.2	26.2	39.4





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.602 min Scan# 649
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	136	Resp:	4144
Ion Ratio	Lower	Upper	
136	100		
137	13.1	9.7	14.5
54	46.9	35.5	53.3
68	7.5	5.0	7.6

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#8
Nitrobenzene-d5
Concen: 0.48 ng
RT: 8.978 min Scan# 524
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Tgt Ion:	82	Resp:	1564
Ion Ratio	Lower	Upper	
82	100		
128	138.8	96.5	144.7
54	293.6	224.7	337.1









#9
Naphthalene
Concen: 0.45 ng
RT: 10.654 min Scan# 653
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	128	Resp:	18533
Ion	Ratio	Lower	Upper
128	100		
129	3.2	2.6	3.8
127	3.8	3.0	4.6

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.927 min Scan# 674
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	225	Resp:	1237
Ion	Ratio	Lower	Upper
225	100		
223	61.9	47.8	71.8
227	65.3	50.5	75.7









#11
2-Methylnaphthalene-d10
Concen: 0.44 ng
RT: 12.210 min Scan# 763
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:152 Resp: 2988
Ion Ratio Lower Upper
152 100
151 18.9 16.2 24.4

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#12
2-Methylnaphthalene
Concen: 0.42 ng
RT: 12.282 min Scan# 768
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41


Tgt Ion:142 Resp: 2871
Ion Ratio Lower Upper
142 100
141 94.7 70.0 105.0
115 45.5 35.2 52.8





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.472 min Scan# 920
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	164	Resp:	2586
Ion	Ratio	Lower	Upper
164	100		
162	99.2	80.5	120.7
160	48.7	40.1	60.1

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.982 min Scan# 1027
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Tgt Ion:	330	Resp:	413
Ion	Ratio	Lower	Upper
330	100		
332	94.4	73.3	109.9
141	46.0	33.1	49.7









#15
2-Fluorobiphenyl
Concen: 0.41 ng
RT: 13.089 min Scan# 824
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	172	Resp:	4368
Ion Ratio	Lower	Upper	
172	100		
171	37.9	29.0	43.4
170	26.0	20.8	31.2

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#16
Acenaphthylene
Concen: 0.45 ng
RT: 14.198 min Scan# 901
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	152	Resp:	5079
Ion Ratio	Lower	Upper	
152	100		
151	21.6	16.3	24.5
153	12.9	10.7	16.1





#17

Acenaphthene

Concen: 0.44 ng

RT: 14.529 min Scan# 924

Delta R.T. 0.000 min

Lab File: BE098757.D

Acq: 26 Feb 2019 12:41

Tgt Ion:	154	Resp:	3072
Ion	Ratio	Lower	Upper
154	100		
153	117.1	93.9	140.9
152	58.5	45.8	68.8

Instrument :





BNA_E

ClientSampled :

SSTDICCC0.4

Manual Integrations
APPROVED

Sohil
2/27/2019 4:22:42 PM



#18

Fluorene

Concen: 0.46 ng




RT: 15.514 min Scan# 992


Delta R.T. 0.000 min

Lab File: BE098757.D

Acq: 26 Feb 2019 12:41

Tgt Ion:	166	Resp:	4175
Ion	Ratio	Lower	Upper
166	100		
165	101.9	79.3	118.9
167	13.7	10.6	15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.227 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:188 Resp: 6289
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.3 4.7 7.1#

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#20
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.411 min Scan# 1059
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41


Tgt Ion:248 Resp: 1313
Ion Ratio Lower Upper
248 100
250 89.5 80.2 120.4
141 96.3 60.5 90.7#





#21
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.531 min Scan# 1068
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tgt Ion:284 Resp: 1700
Ion Ratio Lower Upper
284 100
142 35.5 25.5 38.3
249 34.9 26.6 39.8

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#22
Pentachlorophenol
Concen: 0.27 ng
RT: 16.893 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41




Tgt Ion:266 Resp: 312
Ion Ratio Lower Upper
266 100
264 84.5 53.7 80.5#
268 73.9 59.8 89.8





#23
Phenanthrene
Concen: 0.45 ng
RT: 17.268 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tgt Ion:178 Resp: 6633
Ion Ratio Lower Upper
178 100
176 19.9 15.9 23.9
179 14.9 12.2 18.4

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#24
Anthracene
Concen: 0.41 ng
RT: 17.361 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Tgt Ion:178 Resp: 5189
Ion Ratio Lower Upper
178 100
176 19.4 14.9 22.3
179 16.3 11.8 17.8









#25
Fluoranthene-d10
Concen: 0.45 ng
RT: 19.253 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion: 212 Resp: 34694
Ion Ratio Lower Upper
212 100
106 2.2 1.7 2.5
104 1.2 1.0 1.6

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#26
Fluoranthene
Concen: 0.45 ng
RT: 19.288 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion: 202 Resp: 8536
Ion Ratio Lower Upper
202 100
101 8.0 6.7 10.1
203 16.9 13.3 19.9









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.415 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:240 Resp: 6892
Ion Ratio Lower Upper
240 100
120 6.0 4.7 7.1
236 27.1 21.9 32.9

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#28
Pyrene
Concen: 0.46 ng
RT: 19.650 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41


Tgt Ion:202 Resp: 8683
Ion Ratio Lower Upper
202 100
200 22.3 17.0 25.6
203 17.3 13.8 20.8





#29
Terphenyl-d14
Concen: 0.43 ng
RT: 19.857 min Scan# 1495
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Tgt Ion:244 Resp: 6537
Ion Ratio Lower Upper
244 100
212 37.9 29.4 44.2
122 9.1 7.1 10.7

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#30
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.391 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Tgt Ion:228 Resp: 7261
Ion Ratio Lower Upper
228 100
226 27.3 23.0 34.4
229 20.3 15.0 22.6









#31
Chrysene
Concen: 0.47 ng
RT: 21.451 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	228	Resp:	9014
Ion Ratio	Lower	Upper	
228	100		
226	29.2	24.6	36.8
229	20.1	16.4	24.6

Instrument :
BNA_E
ClientSampled :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.39 ng
RT: 21.306 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	149	Resp:	15353
Ion Ratio	Lower	Upper	
149	100		
167	7.1	5.4	8.2
279	1.1	1.0	1.4









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 26.570 min Scan# 2953
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion:	276	Resp:	7298
Ion Ratio	Lower	Upper	
276	100		
138	17.1	13.3	19.9
277	24.3	19.4	29.0

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


Sohil
2/27/2019 4:22:42 PM



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.921 min Scan# 2214
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41





Tgt Ion:	264	Resp:	6594
Ion Ratio	Lower	Upper	
264	100		
260	25.4	20.5	30.7
265	31.3	22.3	33.5








#35
Benzo(b)fluoranthene
Concen: 0.45 ng m
RT: 23.147 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.5	19.4	29.0
125	13.6	8.2	12.4#



#36
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 23.197 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41


Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.9	19.0	28.4
125	8.7	7.0	10.4




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM



#37
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.807 min Scan# 2182
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41




Tgt Ion:252 Resp: 7455
Ion Ratio Lower Upper
252 100
253 26.2 21.0 31.4
125 15.2 9.2 13.8#




Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED



Sohil
2/27/2019 4:22:42 PM




#38
Dibenzo(a,h)anthracene
Concen: 0.29 ng
RT: 26.595 min Scan# 2960
Delta R.T. 0.000 min
Lab File: BE098757.D
Acq: 26 Feb 2019 12:41



Tgt Ion:278 Resp: 4597
Ion Ratio Lower Upper
278 100
139 18.6 12.7 19.1
279 26.4 22.0 33.0





#39
 Benzo(g,h,i)perylene
 Concen: 0.41 ng
 RT: 27.373 min Scan# 3178
 Delta R.T. 0.000 min
 Lab File: BE098757.D
 Acq: 26 Feb 2019 12:41

Tgt Ion: 276 Resp: 7063
 Ion Ratio Lower Upper
 276 100
 277 25.8 20.3 30.5
 138 16.6 13.3 19.9

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Manual Integrations
 APPROVED

Sohil
 2/27/2019 4:22:42 PM

