

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022719\
 Data File : BE098781.D
 Acq On : 27 Feb 2019 17:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Mar 05 08:19:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 26 15:44:42 2019
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	1014	0.40	ng	0.00
7) Naphthalene-d8	10.602	136	4360	0.40	ng	0.00
13) Acenaphthene-d10	14.472	164	2884	0.40	ng	0.00
19) Phenanthrene-d10	17.227	188	7009	0.40	ng	0.00
27) Chrysene-d12	21.403	240	8269	0.40	ng	#-0.01
34) Perylene-d12	23.920	264	7216	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.400	112	1028	0.39	ng	0.00
5) Phenol-d6	7.000	99	1558	0.39	ng	0.00
8) Nitrobenzene-d5	8.978	82	1681	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	3166	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	518	0.42	ng	-0.01
15) 2-Fluorobiphenyl	13.089	172	4815	0.39	ng	0.00
25) Fluoranthene-d10	19.253	212	37434	0.38	ng	0.00
29) Terphenyl-d14	19.857	244	7222	0.35	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.303	88	672	0.38	ng	94
3) n-Nitrosodimethylamine	3.660	42	703	0.36	ng	# 73
6) bis(2-Chloroethyl)ether	7.242	93	879	0.33	ng	# 81
9) Naphthalene	10.654	128	19917	0.40	ng	100
10) Hexachlorobutadiene	10.927	225	1314	0.39	ng	97
12) 2-Methylnaphthalene	12.282	142	3164	0.39	ng	97
16) Acenaphthylene	14.184	152	5618	0.38	ng	99
17) Acenaphthene	14.529	154	3315	0.38	ng	98
18) Fluorene	15.514	166	4649	0.39	ng	97
20) 4-Bromophenyl-phenylether	16.411	248	1531	0.39	ng	95
21) Hexachlorobenzene	16.531	284	1859	0.39	ng	97
22) Pentachlorophenol	16.893	266	345	0.43	ng	98
23) Phenanthrene	17.268	178	7372	0.38	ng	99
24) Anthracene	17.361	178	5542	0.37	ng	97
26) Fluoranthene	19.288	202	9457	0.38	ng	99
28) Pyrene	19.650	202	9592	0.35	ng	100
30) Benzo(a)anthracene	21.391	228	8514	0.36	ng	98
31) Chrysene	21.451	228	9226	0.34	ng	99
32) Bis(2-ethylhexyl)phtha...	21.306	149	15613	0.31	ng	99
33) Indeno(1,2,3-cd)pyrene	26.577	276	7813	0.37	ng	97
35) Benzo(b)fluoranthene	23.200	252	8437	0.37	ng	96
36) Benzo(k)fluoranthene	23.200	252	8437	0.36	ng	97
37) Benzo(a)pyrene	23.810	252	7681	0.37	ng	# 88
38) Dibenzo(a,h)anthracene	26.591	278	5973	0.44	ng	# 92
39) Benzo(g,h,i)perylene	27.375	276	6807	0.35	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022719\
Data File : BE098781.D
Acq On : 27 Feb 2019 17:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Mar 05 08:19:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 26 15:44:42 2019
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.807 min Scan# 429
 Delta R.T. -0.008 min
 Lab File: BE098781.D
 Acq: 27 Feb 2019 17:46




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:152 Resp: 1014
 Ion Ratio Lower Upper
 152 100
 150 146.7 121.4 182.2
 115 67.8 59.1 88.7




#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.303 min Scan# 38
 Delta R.T. -0.009 min
 Lab File: BE098781.D
 Acq: 27 Feb 2019 17:46

Tgt Ion: 88 Resp: 672
 Ion Ratio Lower Upper
 88 100
 58 100.3 73.9 110.9
 43 52.7 43.7 65.5







Abundance Scan 70 (3.669 min): BE098757.D\data.ms (-64)






#3
n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.660 min Scan# 69
Delta R.T. -0.009 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	42	Resp:	703
Ion	Ratio	Lower	Upper
42	100		
74	50.9	58.9	88.3#
44	0.0	0.0	0.0




#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.400 min Scan# 220
Delta R.T. -0.009 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	112	Resp:	1028
Ion	Ratio	Lower	Upper
112	100		
64	82.4	62.9	94.3
63	188.5	152.2	228.4







Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC



#5
Phenol-d6
Concen: 0.39 ng
RT: 7.000 min Scan# 359
Delta R.T. 0.003 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 1558
Ion Ratio Lower Upper
99 100
42 37.5 24.6 37.0#
71 47.1 35.1 52.7



#6
bis(2-Chloroethyl)ether
Concen: 0.33 ng
RT: 7.242 min Scan# 380
Delta R.T. 0.002 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion: 93 Resp: 879
Ion Ratio Lower Upper
93 100
63 374.4 267.3 400.9
95 40.0 26.2 39.4#









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.602 min Scan# 649
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	4360
Ion Ratio	Lower	Upper	
136	100		
137	12.8	9.7	14.5
54	45.2	35.5	53.3
68	7.0	5.0	7.6



#8
Nitrobenzene-d5
Concen: 0.39 ng
RT: 8.978 min Scan# 524
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	82	Resp:	1681
Ion Ratio	Lower	Upper	
82	100		
128	150.3	96.5	144.7#
54	278.7	224.7	337.1









#9
Naphthalene
Concen: 0.40 ng
RT: 10.654 min Scan# 653
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	19917
Ion	Ratio	Lower	Upper
128	100		
129	3.3	2.6	3.8
127	3.8	3.0	4.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.927 min Scan# 674
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	225	Resp:	1314
Ion	Ratio	Lower	Upper
225	100		
223	62.3	47.8	71.8
227	61.6	50.5	75.7









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.210 min Scan# 763
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 3166
Ion Ratio Lower Upper
152 100
151 18.1 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.282 min Scan# 768
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:142 Resp: 3164
Ion Ratio Lower Upper
142 100
141 85.3 70.0 105.0
115 46.7 35.2 52.8









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.472 min Scan# 920
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164	Resp:	2884	
Ion	Ratio	Lower	Upper
164	100		
162	98.0	80.5	120.7
160	50.1	40.1	60.1



#14
2,4,6-Tribromophenol
Concen: 0.42 ng
RT: 15.969 min Scan# 1026
Delta R.T. -0.013 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:330	Resp:	518	
Ion	Ratio	Lower	Upper
330	100		
332	97.7	73.3	109.9
141	40.0	33.1	49.7





#15

2-Fluorobiphenyl

Concen: 0.39 ng

RT: 13.089 min Scan# 824

Delta R.T. -0.000 min

Lab File: BE098781.D

Acq: 27 Feb 2019 17:46





Tgt Ion:	172	Resp:	4815
Ion	Ratio	Lower	Upper
172	100		
171	35.4	29.0	43.4
170	25.2	20.8	31.2

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC



#16

Acenaphthylene

Concen: 0.38 ng




RT: 14.184 min Scan# 900


Delta R.T. -0.014 min

Lab File: BE098781.D

Acq: 27 Feb 2019 17:46

Tgt Ion:	152	Resp:	5618
Ion	Ratio	Lower	Upper
152	100		
151	20.7	16.3	24.5
153	13.1	10.7	16.1









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.529 min Scan# 924
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	154	Resp:	3315
Ion Ratio	Lower	Upper	
154	100		
153	119.1	93.9	140.9
152	60.1	45.8	68.8



#18
Fluorene
Concen: 0.39 ng
RT: 15.514 min Scan# 992
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	166	Resp:	4649
Ion Ratio	Lower	Upper	
166	100		
165	101.6	79.3	118.9
167	14.3	10.6	15.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.227 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 7009
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 4.7 7.1



#20
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.411 min Scan# 1059
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:248 Resp: 1531
Ion Ratio Lower Upper
248 100
250 91.7 80.2 120.4
141 74.6 60.5 90.7









#21
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.531 min Scan# 1068
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:284 Resp: 1859
Ion Ratio Lower Upper
284 100
142 34.2 25.5 38.3
249 34.4 26.6 39.8



#22
Pentachlorophenol
Concen: 0.43 ng
RT: 16.893 min Scan# 1095
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:266 Resp: 345
Ion Ratio Lower Upper
266 100
264 69.8 53.7 80.5
268 75.4 59.8 89.8









#23
Phenanthrene
Concen: 0.38 ng
RT: 17.268 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 7372
Ion Ratio Lower Upper
178 100
176 19.5 15.9 23.9
179 16.0 12.2 18.4



#24
Anthracene
Concen: 0.37 ng
RT: 17.361 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:178 Resp: 5542
Ion Ratio Lower Upper
178 100
176 20.0 14.9 22.3
179 15.9 11.8 17.8









#25
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.253 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 37434
Ion Ratio Lower Upper
212 100
106 2.2 1.7 2.5
104 1.2 1.0 1.6



#26
Fluoranthene
Concen: 0.38 ng
RT: 19.288 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:202 Resp: 9457
Ion Ratio Lower Upper
202 100
101 8.1 6.7 10.1
203 16.9 13.3 19.9









#27
Chrysene-d12
Concen: 0.40 ng
RT: 21.403 min Scan# 1633
Delta R.T. -0.012 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 8269
Ion Ratio Lower Upper
240 100
120 11.1 4.7 7.1#
236 29.2 21.9 32.9



#28
Pyrene
Concen: 0.35 ng
RT: 19.650 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:202 Resp: 9592
Ion Ratio Lower Upper
202 100
200 21.5 17.0 25.6
203 17.0 13.8 20.8





#29

Terphenyl-d14

Concen: 0.35 ng

RT: 19.857 min Scan# 1495

Delta R.T. 0.000 min

Lab File: BE098781.D

Acq: 27 Feb 2019 17:46





Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:244	Resp:	7222
Ion Ratio	Lower	Upper
244	100	
212	36.8	29.4 44.2
122	9.3	7.1 10.7



#30

Benzo(a)anthracene

Concen: 0.36 ng




RT: 21.391 min Scan# 1632


Delta R.T. -0.000 min

Lab File: BE098781.D

Acq: 27 Feb 2019 17:46

Tgt Ion:228	Resp:	8514
Ion Ratio	Lower	Upper
228	100	
226	28.8	23.0 34.4
229	20.9	15.0 22.6









#31
Chrysene
Concen: 0.34 ng
RT: 21.451 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	9226
Ion Ratio	Lower	Upper	
228	100		
226	30.1	24.6	36.8
229	20.8	16.4	24.6



#32
Bis(2-ethylhexyl)phthalate
Concen: 0.31 ng
RT: 21.306 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:	149	Resp:	15613
Ion Ratio	Lower	Upper	
149	100		
167	7.0	5.4	8.2
279	1.1	1.0	1.4









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 26.577 min Scan# 2955
Delta R.T. 0.007 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 7813
Ion Ratio Lower Upper
276 100
138 19.8 13.3 19.9
277 24.6 19.4 29.0



#34
Perylene-d12
Concen: 0.40 ng
RT: 23.920 min Scan# 2214
Delta R.T. -0.001 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:264 Resp: 7216
Ion Ratio Lower Upper
264 100
260 25.6 20.5 30.7
265 28.6 22.3 33.5









#35
Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 23.200 min Scan# 2012
Delta R.T. 0.053 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 8437
Ion Ratio Lower Upper
252 100
253 25.8 19.4 29.0
125 8.6 8.2 12.4



#36
Benzo(k)fluoranthene
Concen: 0.36 ng
RT: 23.200 min Scan# 2012
Delta R.T. 0.003 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:252 Resp: 8437
Ion Ratio Lower Upper
252 100
253 25.8 19.0 28.4
125 8.6 7.0 10.4









#37
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.810 min Scan# 2183
Delta R.T. 0.003 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 7681
Ion Ratio Lower Upper
252 100
253 27.6 21.0 31.4
125 24.2 9.2 13.8#



#38
Dibenzo(a,h)anthracene
Concen: 0.44 ng
RT: 26.591 min Scan# 2959
Delta R.T. -0.004 min
Lab File: BE098781.D
Acq: 27 Feb 2019 17:46

Tgt Ion:278 Resp: 5973
Ion Ratio Lower Upper
278 100
139 23.6 12.7 19.1#
279 26.5 22.0 33.0





#39
 Benzo(g,h,i)perylene
 Concen: 0.35 ng
 RT: 27.375 min Scan# 3179
 Delta R.T. 0.003 min
 Lab File: BE098781.D
 Acq: 27 Feb 2019 17:46

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	6807
Ion Ratio	Lower	Upper	
276	100		
277	25.9	20.3	30.5
138	17.4	13.3	19.9

