

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE022719\
 Data File : BE098771.D
 Acq On : 27 Feb 2019 9:55
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 27 11:27:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE022619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 26 15:44:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.690	0.595	13.8	88	0.00
3	n-Nitrosodimethylamine	0.758	0.655	13.6	86	0.00
4 S	2-Fluorophenol	1.043	1.018	2.4	101	0.00
5 S	Phenol-d6	1.560	1.545	1.0	103	0.00
6	bis(2-Chloroethyl)ether	1.071	1.013	5.4	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.380	0.396	-4.2	108	0.00
9	Naphthalene	4.611	4.513	2.1	103	0.00
10	Hexachlorobutadiene	0.306	0.287	6.2	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.757	0.733	3.2	104	0.00
12	2-Methylnaphthalene	0.735	0.721	1.9	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.158	0.172	-8.9	118	0.00
15 S	2-Fluorobiphenyl	1.693	1.634	3.5	106	0.00
16	Acenaphthylene	2.045	1.961	4.1	110	0.00
17	Acenaphthene	1.209	1.156	4.4	107	0.00
18	Fluorene	1.673	1.621	3.1	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4-Bromophenyl-phenylether	0.225	0.225	0.0	114	0.00
21	Hexachlorobenzene	0.270	0.269	0.4	105	0.00
22	Pentachlorophenol	0.061	0.069	-13.1	148	0.00
23	Phenanthrene	1.096	1.061	3.2	107	0.00
24	Anthracene	0.885	0.864	2.4	111	0.00
25 SURR	Fluoranthene-d10	5.692	5.676	0.3	109	0.00
26	Fluoranthene	1.414	1.394	1.4	109	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
28	Pyrene	1.329	1.340	-0.8	111	0.00
29 S	Terphenyl-d14	0.992	0.990	0.2	109	0.00
30	Benzo(a)anthracene	1.152	1.228	-6.6	122	0.00
31	Chrysene	1.300	1.252	3.7	100	0.00
32	Bis(2-ethylhexyl)phthalate	2.456	2.676	-9.0	126	0.00
33	Indeno(1,2,3-cd)pyrene	1.032	1.069	-3.6	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.260	1.244	1.3	97	0.00
36	Benzo(k)fluoranthene	1.313	1.339	-2.0	101	0.00
37 C	Benzo(a)pyrene	1.163	1.187	-2.1	102	0.00
38	Dibenzo(a,h)anthracene	0.739	0.892	-20.7	124	0.00
39	Benzo(g,h,i)perylene	1.085	1.075	0.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0