

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030619\
 Data File : BE098805.D
 Acq On : 6 Mar 2019 19:26
 Operator : JU/SJ
 Sample : K1659-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 M1-T4A-UV(POST)

Quant Time: Mar 07 02:44:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	783	0.40	ng	0.00
7) Naphthalene-d8	10.59	136	3300	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	2146	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	5813	0.40	ng	0.00
27) Chrysene-d12	21.40	240	7523	0.40	ng	0.00
34) Perylene-d12	23.90	264	7436	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	328	0.14	ng	0.00
5) Phenol-d6	6.99	99	286	0.09	ng	0.00
8) Nitrobenzene-d5	8.98	82	985	0.28	ng	0.01
11) 2-Methylnaphthalene-d10	12.19	152	2075	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	390	0.36	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	5360	0.60	ng	0.00
25) Fluoranthene-d10	19.25	212	32574	0.35	ng	0.00
29) Terphenyl-d14	19.84	244	7862	0.43	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.31	149	104647	2.094	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE030619\
Data File : BE098805.D
Acq On : 6 Mar 2019 19:26
Operator : JU/SJ
Sample : K1659-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
M1-T4A-UV(POST)

Quant Time: Mar 07 02:44:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

