

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE030619\
 Data File : BE098808.D
 Acq On : 6 Mar 2019 21:16
 Operator : JU/SJ
 Sample : K1655-05
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A0C1-TWP

Quant Time: Mar 12 12:28:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	872	0.40	ng	0.00
7) Naphthalene-d8	10.59	136	3529	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	2318	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	6146	0.40	ng	0.00
27) Chrysene-d12	21.40	240	7778	0.40	ng	0.00
34) Perylene-d12	23.90	264	7646	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	312	0.12	ng	0.00
5) Phenol-d6	6.98	99	334	0.09	ng	0.00
8) Nitrobenzene-d5	8.96	82	1145	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.19	152	2520	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.96	330	419	0.36	ng	-0.01
15) 2-Fluorobiphenyl	13.07	172	5041	0.53	ng	-0.01
25) Fluoranthene-d10	19.24	212	39310	0.40	ng	0.00
29) Terphenyl-d14	19.84	244	9995	0.53	ng	0.00
Target Compounds						
9) Naphthalene	10.64	128	1626	0.040	ng	# 84
12) 2-Methylnaphthalene	12.27	142	303	0.046	ng	# 82
32) Bis(2-ethylhexyl)phthalate	21.30	149	4196	0.081	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE030619\
Data File : BE098808.D
Acq On : 6 Mar 2019 21:16
Operator : JU/SJ
Sample : K1655-05
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A0C1-TWP

Quant Time: Mar 12 12:28:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

