

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE030819\  
 Data File : BE098831.D  
 Acq On : 8 Mar 2019 10:42  
 Operator : JU/SJ  
 Sample : PB117716BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 PB117716BL

Quant Time: Mar 08 13:41:26 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE030619.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 06 15:33:24 2019  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.79  | 152  | 649      | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8         | 10.60 | 136  | 2675     | 0.40 | ng    | 0.01     |
| 13) Acenaphthene-d10      | 14.47 | 164  | 1973     | 0.40 | ng    | 0.01     |
| 19) Phenanthrene-d10      | 17.22 | 188  | 5303     | 0.40 | ng    | 0.00     |
| 27) Chrysene-d12          | 21.40 | 240  | 6373     | 0.40 | ng    | 0.00     |
| 34) Perylene-d12          | 23.90 | 264  | 6438     | 0.40 | ng    | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 4) 2-Fluorophenol           | 5.40  | 112  | 448      | 0.23 | ng    | 0.01     |
| 5) Phenol-d6                | 7.02  | 99   | 571      | 0.21 | ng    | 0.04     |
| 8) Nitrobenzene-d5          | 9.03  | 82   | 333      | 0.12 | ng    | 0.07     |
| 11) 2-Methylnaphthalene-d10 | 12.22 | 152  | 1143     | 0.23 | ng    | 0.03     |
| 14) 2,4,6-Tribromophenol    | 15.98 | 330  | 158      | 0.16 | ng    | 0.01     |
| 15) 2-Fluorobiphenyl        | 13.10 | 172  | 1773     | 0.22 | ng    | 0.01     |
| 25) Fluoranthene-d10        | 19.25 | 212  | 18539    | 0.22 | ng    | 0.00     |
| 29) Terphenyl-d14           | 19.85 | 244  | 3724     | 0.24 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE030819\  
Data File : BE098831.D  
Acq On : 8 Mar 2019 10:42  
Operator : JU/SJ  
Sample : PB117716BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
PB117716BL

Quant Time: Mar 08 13:41:26 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE030619.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 06 15:33:24 2019  
Response via : Initial Calibration

