

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE030819\
 Data File : BE098838.D
 Acq On : 8 Mar 2019 17:25
 Operator : JU/SJ
 Sample : K1793-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE131D1-20190305

Quant Time: Mar 09 01:50:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	737	0.40	ng	0.00
7) Naphthalene-d8	10.59	136	2986	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	2142	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	5195	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6522	0.40	ng	0.00
34) Perylene-d12	23.90	264	6441	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.40	112	326	0.15	ng	0.01
5) Phenol-d6	7.01	99	226	0.07	ng	0.02
8) Nitrobenzene-d5	8.98	82	1224	0.39	ng	0.01
11) 2-Methylnaphthalene-d10	12.19	152	2218	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	416	0.38	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	3645	0.41	ng	0.00
25) Fluoranthene-d10	19.25	212	33383	0.40	ng	0.00
29) Terphenyl-d14	19.84	244	8260	0.52	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.28	88	10014	8.721	ng	99
32) Bis(2-ethylhexyl)phthalate	21.31	149	1591	0.037	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE030819\
Data File : BE098838.D
Acq On : 8 Mar 2019 17:25
Operator : JU/SJ
Sample : K1793-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE131D1-20190305

Quant Time: Mar 09 01:50:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

