

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE030819\
 Data File : BE098841.D
 Acq On : 8 Mar 2019 19:15
 Operator : JU/SJ
 Sample : K1793-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DUP01-20190305

Quant Time: Mar 09 01:52:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	804	0.40	ng	0.00
7) Naphthalene-d8	10.59	136	3130	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	2140	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	5283	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6578	0.40	ng	0.00
34) Perylene-d12	23.89	264	6574	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.40	112	443	0.18	ng	0.01
5) Phenol-d6	7.00	99	301	0.09	ng	0.01
8) Nitrobenzene-d5	8.98	82	1332	0.41	ng	0.01
11) 2-Methylnaphthalene-d10	12.20	152	2350	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	499	0.46	ng	0.00
15) 2-Fluorobiphenyl	13.07	172	4190	0.47	ng	-0.01
25) Fluoranthene-d10	19.24	212	36702	0.44	ng	0.00
29) Terphenyl-d14	19.84	244	9392	0.59	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.28	88	11306	9.030	ng	100
32) Bis(2-ethylhexyl)phthalate	21.29	149	2825	0.065	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE030819\
Data File : BE098841.D
Acq On : 8 Mar 2019 19:15
Operator : JU/SJ
Sample : K1793-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DUP01-20190305

Quant Time: Mar 09 01:52:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

