

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE030819\
 Data File : BE098844.D
 Acq On : 8 Mar 2019 22:17
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 09 05:07:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.789	0.749	5.1	119	0.00
3	n-Nitrosodimethylamine	1.006	0.857	14.8	101	0.01
4 S	2-Fluorophenol	1.196	1.123	6.1	109	0.01
5 S	Phenol-d6	1.689	1.624	3.8	116	0.00
6	bis(2-Chloroethyl)ether	1.304	1.023	21.5	98	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.419	0.376	10.3	125	0.00
9	Naphthalene	4.581	4.549	0.7	128	-0.01
10	Hexachlorobutadiene	0.303	0.296	2.3	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.732	0.9	127	0.00
12	2-Methylnaphthalene	0.743	0.721	3.0	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.202	0.197	2.5	130	-0.01
15 S	2-Fluorobiphenyl	1.656	1.584	4.3	128	-0.01
16	Acenaphthylene	2.057	1.966	4.4	133	0.00
17	Acenaphthene	1.220	1.187	2.7	129	-0.01
18	Fluorene	1.760	1.685	4.3	130	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	-0.01
20	4-Bromophenyl-phenylether	0.217	0.197	9.2	119	0.00
21	Hexachlorobenzene	0.258	0.237	8.1	125	0.00
22	Pentachlorophenol	0.052	0.034	34.6#	114	0.00
23	Phenanthrene	1.137	1.069	6.0	124	-0.01
24	Anthracene	0.972	0.854	12.1	120	0.00
25 SURR	Fluoranthene-d10	6.385	5.588	12.5	117	0.00
26	Fluoranthene	1.605	1.424	11.3	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	-0.01
28	Pyrene	1.224	1.217	0.6	119	0.00
29 S	Terphenyl-d14	0.972	0.923	5.0	114	0.00
30	Benzo(a)anthracene	1.248	1.109	11.1	108	0.00
31	Chrysene	1.336	1.314	1.6	120	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	2.235	15.9	110	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.378	2.3	119	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	125	0.00
35	Benzo(b)fluoranthene	1.315	1.254	4.6	123	0.00
36	Benzo(k)fluoranthene	1.377	1.306	5.2	124	-0.01
37 C	Benzo(a)pyrene	1.261	1.220	3.3	124	-0.01
38	Dibenzo(a,h)anthracene	1.193	1.132	5.1	120	0.00
39	Benzo(g,h,i)perylene	1.248	1.187	4.9	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE030819\
Data File : BE098844.D
Acq On : 8 Mar 2019 22:17
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 09 05:07:39 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			