

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE030819\
 Data File : BE098855.D
 Acq On : 9 Mar 2019 5:12
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 09 06:59:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	128	0.00
2	1,4-Dioxane	0.400	0.400	0.0	131	0.00
3	n-Nitrosodimethylamine	0.400	0.288	28.0#	90	0.00
4 S	2-Fluorophenol	0.400	0.343	14.2	106	0.00
5 S	Phenol-d6	0.400	0.370	7.5	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.364	9.0	121	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	135	0.00
8 S	Nitrobenzene-d5	0.400	0.341	14.8	128	0.00
9	Naphthalene	0.400	0.390	2.5	135	-0.01
10	Hexachlorobutadiene	0.400	0.389	2.8	135	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	133	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	138	0.00
14 S	2,4,6-Tribromophenol	0.400	0.377	5.8	132	-0.01
15 S	2-Fluorobiphenyl	0.400	0.398	0.5	139	-0.01
16	Acenaphthylene	0.400	0.388	3.0	141	0.00
17	Acenaphthene	0.400	0.391	2.3	136	-0.01
18	Fluorene	0.400	0.383	4.3	136	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	134	-0.01
20	4-Bromophenyl-phenylether	0.400	0.363	9.3	123	0.00
21	Hexachlorobenzene	0.400	0.382	4.5	134	0.00
22	Pentachlorophenol	0.400	0.427	-6.7	124	0.00
23	Phenanthrene	0.400	0.373	6.8	128	0.00
24	Anthracene	0.400	0.359	10.3	127	0.00
25 SURR	Fluoranthene-d10	0.400	0.352	12.0	122	0.00
26	Fluoranthene	0.400	0.351	12.3	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	129	-0.01
28	Pyrene	0.400	0.383	4.3	123	0.00
29 S	Terphenyl-d14	0.400	0.371	7.3	120	0.00
30	Benzo(a)anthracene	0.400	0.360	10.0	118	0.00
31	Chrysene	0.400	0.388	3.0	128	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.330	17.5	116	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	131	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	134	-0.01
35	Benzo(b)fluoranthene	0.400	0.383	4.3	133	0.00
36	Benzo(k)fluoranthene	0.400	0.395	1.3	139	-0.01
37 C	Benzo(a)pyrene	0.400	0.386	3.5	134	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.390	2.5	133	-0.02
39	Benzo(g,h,i)perylene	0.400	0.382	4.5	131	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE030819\
Data File : BE098855.D
Acq On : 9 Mar 2019 5:12
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 09 06:59:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				