

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098864.D
 Acq On : 11 Mar 2019 15:29
 Operator : JU/SJ
 Sample : K1793-04DL 5X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DUP01-20190305DL

Manual Integrations
 APPROVED

Sohil
 3/12/2019 4:33:49 PM

Quant Time: Mar 12 08:11:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	636	0.40	ng	0.01
7) Naphthalene-d8	10.60	136	3062	0.40	ng	0.01
13) Acenaphthene-d10	14.47	164	2271	0.40	ng	0.01
19) Phenanthrene-d10	17.21	188	5934	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6463	0.40	ng	0.00
34) Perylene-d12	23.90	264	6497	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.43	112	44	0.02	ng	0.05
5) Phenol-d6	7.11	99	52m	0.02	ng	0.13
8) Nitrobenzene-d5	8.96	82	16	0.00	ng	0.00
11) 2-Methylnaphthalene-d10	12.24	152	454	0.08	ng	0.04
14) 2,4,6-Tribromophenol	16.00	330	63	0.05	ng	0.03
15) 2-Fluorobiphenyl	13.12	172	851	0.09	ng	0.03
25) Fluoranthene-d10	19.25	212	7371	0.08	ng	0.00
29) Terphenyl-d14	19.85	244	1717	0.11	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.28	88	2372	2.318	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098864.D
Acq On : 11 Mar 2019 15:29
Operator : JU/SJ
Sample : K1793-04DL 5X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DUP01-20190305DL

Manual Integrations
APPROVED

Sohil
3/12/2019 4:33:49 PM

Quant Time: Mar 12 08:11:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

