

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098866.D
 Acq On : 11 Mar 2019 17:40
 Operator : JU/SJ
 Sample : K1839-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE109D2-20190307

Quant Time: Mar 12 07:29:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	815	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3247	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2146	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	6123	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6895	0.40	ng	0.00
34) Perylene-d12	23.91	264	4711	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	220	0.09	ng	0.04
5) Phenol-d6	6.97	99	274	0.08	ng	-0.01
8) Nitrobenzene-d5	9.00	82	796	0.23	ng	0.04
11) 2-Methylnaphthalene-d10	12.21	152	1704	0.28	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	316	0.29	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	2880	0.32	ng	0.00
25) Fluoranthene-d10	19.25	212	25079	0.26	ng	0.00
29) Terphenyl-d14	19.84	244	6988	0.42	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.28	88	4360	3.371	ng	Qvalue 98
32) Bis(2-ethylhexyl)phthalate	21.29	149	4008	0.088	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098866.D
Acq On : 11 Mar 2019 17:40
Operator : JU/SJ
Sample : K1839-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE109D2-20190307

Quant Time: Mar 12 07:29:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

