

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098880.D
 Acq On : 12 Mar 2019 3:31
 Operator : JU/SJ
 Sample : K1841-05
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE122D3-20190308

Quant Time: Mar 12 07:49:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	735	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3122	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2228	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	5826	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6855	0.40	ng	0.00
34) Perylene-d12	23.90	264	6589	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	254	0.12	ng	0.03
5) Phenol-d6	7.01	99	238	0.08	ng	0.03
8) Nitrobenzene-d5	9.00	82	1018	0.31	ng	0.04
11) 2-Methylnaphthalene-d10	12.21	152	2031	0.35	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	359	0.32	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	3377	0.37	ng	0.00
25) Fluoranthene-d10	19.25	212	32368	0.35	ng	0.00
29) Terphenyl-d14	19.84	244	8086	0.49	ng	0.00
Target Compounds						
32) Bis(2-ethylhexyl)phthalate	21.29	149	5529	0.121	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098880.D
Acq On : 12 Mar 2019 3:31
Operator : JU/SJ
Sample : K1841-05
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE122D3-20190308

Quant Time: Mar 12 07:49:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

