

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098883.D
 Acq On : 12 Mar 2019 5:22
 Operator : JU/SJ
 Sample : K1841-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE125D3-20190308

Quant Time: Mar 12 07:51:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	832	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3529	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2539	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	6572	0.40	ng	0.00
27) Chrysene-d12	21.40	240	7830	0.40	ng	0.00
34) Perylene-d12	23.90	264	7485	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	359	0.14	ng	0.03
5) Phenol-d6	7.03	99	180	0.05	ng	0.05
8) Nitrobenzene-d5	8.99	82	1167	0.32	ng	0.03
11) 2-Methylnaphthalene-d10	12.21	152	2433	0.37	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	394	0.31	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4060	0.39	ng	0.00
25) Fluoranthene-d10	19.24	212	35797	0.34	ng	0.00
29) Terphenyl-d14	19.84	244	9035	0.47	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.28	88	4136	3.125	ng	96
32) Bis(2-ethylhexyl)phthalate	21.29	149	1424	0.027	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098883.D
Acq On : 12 Mar 2019 5:22
Operator : JU/SJ
Sample : K1841-08
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE125D3-20190308

Quant Time: Mar 12 07:51:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

