

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098857.D
 Acq On : 11 Mar 2019 11:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 12 16:30:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.789	0.779	1.3	113	0.00
3	n-Nitrosodimethylamine	1.006	0.560	44.3#	60	0.02
4 S	2-Fluorophenol	1.196	0.936	21.7	83	0.01
5 S	Phenol-d6	1.689	1.566	7.3	102	0.00
6	bis(2-Chloroethyl)ether	1.304	1.112	14.7	97	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.419	0.365	12.9	110	0.00
9	Naphthalene	4.581	4.617	-0.8	118	0.00
10	Hexachlorobutadiene	0.303	0.291	4.0	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.725	1.9	114	0.00
12	2-Methylnaphthalene	0.743	0.711	4.3	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.202	0.161	20.3	95	0.00
15 S	2-Fluorobiphenyl	1.656	1.702	-2.8	123	0.00
16	Acenaphthylene	2.057	1.973	4.1	119	0.00
17	Acenaphthene	1.220	1.170	4.1	114	0.00
18	Fluorene	1.760	1.663	5.5	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.217	0.195	10.1	100	0.00
21	Hexachlorobenzene	0.258	0.251	2.7	112	0.00
22	Pentachlorophenol	0.052	0.026	50.0#	73	0.01
23	Phenanthrene	1.137	1.080	5.0	107	0.00
24	Anthracene	0.972	0.863	11.2	103	0.00
25 SURR	Fluoranthene-d10	6.385	5.654	11.4	101	0.00
26	Fluoranthene	1.605	1.437	10.5	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.224	1.244	-1.6	101	0.00
29 S	Terphenyl-d14	0.972	0.930	4.3	95	0.00
30	Benzo(a)anthracene	1.248	1.193	4.4	96	0.00
31	Chrysene	1.336	1.287	3.7	98	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	2.051	22.8	83	0.00
33	Indeno(1,2,3-cd)pyrene	1.410	1.400	0.7	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.315	1.260	4.2	99	0.00
36	Benzo(k)fluoranthene	1.377	1.407	-2.2	108	0.00
37 C	Benzo(a)pyrene	1.261	1.226	2.8	101	0.00
38	Dibenzo(a,h)anthracene	1.193	1.123	5.9	96	0.00
39	Benzo(a,h,i)perylene	1.248	1.230	1.4	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098857.D
Acq On : 11 Mar 2019 11:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 12 16:30:39 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			