

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
 Data File : BE098888.D
 Acq On : 12 Mar 2019 8:24
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 12 16:39:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.789	0.654	17.1	99	0.00
3	n-Nitrosodimethylamine	1.006	0.448	55.5#	50	0.01
4 S	2-Fluorophenol	1.196	0.984	17.7	91	0.01
5 S	Phenol-d6	1.689	1.560	7.6	106	0.00
6	bis(2-Chloroethyl)ether	1.304	1.048	19.6	96	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.419	0.352	16.0	111	0.00
9	Naphthalene	4.581	4.569	0.3	122	-0.01
10	Hexachlorobutadiene	0.303	0.289	4.6	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.737	0.3	121	0.00
12	2-Methylnaphthalene	0.743	0.722	2.8	120	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.202	0.166	17.8	105	-0.01
15 S	2-Fluorobiphenyl	1.656	1.616	2.4	125	0.00
16	Acenaphthylene	2.057	1.901	7.6	123	0.00
17	Acenaphthene	1.220	1.164	4.6	122	-0.01
18	Fluorene	1.760	1.634	7.2	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.217	0.197	9.2	107	0.00
21	Hexachlorobenzene	0.258	0.256	0.8	121	0.00
22	Pentachlorophenol	0.052	0.037	28.8#	112	0.00
23	Phenanthrene	1.137	1.092	4.0	114	0.00
24	Anthracene	0.972	0.913	6.1	115	0.00
25 SURR	Fluoranthene-d10	6.385	5.765	9.7	108	0.00
26	Fluoranthene	1.605	1.445	10.0	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	239#	-0.01
28	Pvrene	1.224	0.557	54.5#	109	0.00
29 S	Terphenyl-d14	0.972	0.813	16.4	202#	0.00
30	Benzo(a)anthracene	1.248	0.955	23.5	187#	0.00
31	Chrysene	1.336	1.145	14.3	210#	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.166	18.5	213#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.224	13.2	212#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	236#	-0.01
35	Benzo(b)fluoranthene	1.315	1.177	10.5	218#	0.00
36	Benzo(k)fluoranthene	1.377	1.301	5.5	234#	-0.01
37 C	Benzo(a)pyrene	1.261	1.067	15.4	206#	-0.02
38	Dibenzo(a,h)anthracene	1.193	1.083	9.2	217#	-0.02
39	Benzo(a,h,i)perylene	1.248	1.123	10.0	217#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031119\
Data File : BE098888.D
Acq On : 12 Mar 2019 8:24
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 12 16:39:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			