

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
 Data File : BE098892.D
 Acq On : 12 Mar 2019 11:27
 Operator : JU/SJ
 Sample : PB117818BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB117818BL

Quant Time: Mar 12 12:34:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	603	0.40	ng	0.01
7) Naphthalene-d8	10.61	136	2486	0.40	ng	0.02
13) Acenaphthene-d10	14.47	164	1866	0.40	ng	0.02
19) Phenanthrene-d10	17.23	188	4983	0.40	ng	0.01
27) Chrysene-d12	21.40	240	6102	0.40	ng	0.00
34) Perylene-d12	23.90	264	5268	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.40	112	613	0.34	ng	0.01
5) Phenol-d6	7.04	99	730	0.29	ng	0.06
8) Nitrobenzene-d5	9.05	82	645	0.25	ng	0.09
11) 2-Methylnaphthalene-d10	12.22	152	1389	0.30	ng	0.03
14) 2,4,6-Tribromophenol	15.98	330	190	0.20	ng	0.01
15) 2-Fluorobiphenyl	13.10	172	2448	0.32	ng	0.02
25) Fluoranthene-d10	19.25	212	23548	0.30	ng	0.00
29) Terphenyl-d14	19.85	244	4693	0.32	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
Data File : BE098892.D
Acq On : 12 Mar 2019 11:27
Operator : JU/SJ
Sample : PB117818BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB117818BL

Quant Time: Mar 12 12:34:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

