

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
 Data File : BE098907.D
 Acq On : 12 Mar 2019 20:45
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 13 08:10:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.789	0.733	7.1	110	0.00
3	n-Nitrosodimethylamine	1.006	0.811	19.4	90	0.04
4 S	2-Fluorophenol	1.196	0.995	16.8	91	0.01
5 S	Phenol-d6	1.689	1.610	4.7	108	0.00
6	bis(2-Chloroethyl)ether	1.304	1.085	16.8	98	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.419	0.324	22.7	108	0.00
9	Naphthalene	4.581	4.531	1.1	127	0.00
10	Hexachlorobutadiene	0.303	0.304	-0.3	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.713	3.5	123	0.00
12	2-Methylnaphthalene	0.743	0.722	2.8	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.202	0.149	26.2#	96	0.00
15 S	2-Fluorobiphenyl	1.656	1.632	1.4	128	0.00
16	Acenaphthylene	2.057	1.962	4.6	129	0.00
17	Acenaphthene	1.220	1.214	0.5	129	-0.01
18	Fluorene	1.760	1.695	3.7	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.217	0.195	10.1	111	0.00
21	Hexachlorobenzene	0.258	0.250	3.1	124	0.00
22	Pentachlorophenol	0.052	0.026	50.0#	81	0.01
23	Phenanthrene	1.137	1.076	5.4	117	0.00
24	Anthracene	0.972	0.797	18.0	105	0.00
25 SURR	Fluoranthene-d10	6.385	5.438	14.8	107	0.00
26	Fluoranthene	1.605	1.382	13.9	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pvrene	1.224	1.236	-1.0	105	0.00
29 S	Terphenyl-d14	0.972	0.950	2.3	102	0.00
30	Benzo(a)anthracene	1.248	1.047	16.1	89	0.00
31	Chrysene	1.336	1.373	-2.8	110	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	1.864	29.8#	80	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.430	-1.4	107	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	109	-0.01
35	Benzo(b)fluoranthene	1.315	1.256	4.5	107	0.00
36	Benzo(k)fluoranthene	1.377	1.404	-2.0	116	0.00
37 C	Benzo(a)pyrene	1.261	1.228	2.6	109	0.00
38	Dibenzo(a,h)anthracene	1.193	1.118	6.3	103	-0.01
39	Benzo(a,h,i)perylene	1.248	1.216	2.6	108	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
Data File : BE098907.D
Acq On : 12 Mar 2019 20:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 13 08:10:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			