

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
 Data File : BE098907.D
 Acq On : 12 Mar 2019 20:45
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 13 08:10:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 15:33:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.372	7.0	110	0.00
3	n-Nitrosodimethylamine	0.400	0.323	19.3	90	0.04
4 S	2-Fluorophenol	0.400	0.333	16.8	91	0.01
5 S	Phenol-d6	0.400	0.381	4.8	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.333	16.8	98	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	125	0.00
8 S	Nitrobenzene-d5	0.400	0.309	22.8	108	0.00
9	Naphthalene	0.400	0.396	1.0	127	0.00
10	Hexachlorobutadiene	0.400	0.401	-0.3	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	123	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	126	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.400	0.296	26.0#	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	128	0.00
16	Acenaphthylene	0.400	0.382	4.5	129	0.00
17	Acenaphthene	0.400	0.398	0.5	129	-0.01
18	Fluorene	0.400	0.385	3.8	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.400	0.360	10.0	111	0.00
21	Hexachlorobenzene	0.400	0.388	3.0	124	0.00
22	Pentachlorophenol	0.400	0.382	4.5	81	0.01
23	Phenanthrene	0.400	0.379	5.3	117	0.00
24	Anthracene	0.400	0.328	18.0	105	0.00
25 SURR	Fluoranthene-d10	0.400	0.341	14.8	107	0.00
26	Fluoranthene	0.400	0.344	14.0	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	-0.01
28	Pvrene	0.400	0.404	-1.0	105	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	102	0.00
30	Benzo(a)anthracene	0.400	0.335	16.3	89	0.00
31	Chrysene	0.400	0.411	-2.7	110	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.281	29.8#	80	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	107	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	109	-0.01
35	Benzo(b)fluoranthene	0.400	0.382	4.5	107	0.00
36	Benzo(k)fluoranthene	0.400	0.408	-2.0	116	0.00
37 C	Benzo(a)pyrene	0.400	0.389	2.8	109	0.00
38	Dibenzo(a,h)anthracene	0.400	0.375	6.3	103	-0.01
39	Benzo(a,h,i)perylene	0.400	0.390	2.5	108	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031219\
Data File : BE098907.D
Acq On : 12 Mar 2019 20:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 13 08:10:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 15:33:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			