

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
 Data File : BE098991.D
 Acq On : 16 Mar 2019 6:47
 Operator : JU/SJ
 Sample : k1912-13RE
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 M3-T4-UV(POST-HYPO)RE

Quant Time: Mar 18 02:11:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	830	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3812	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2784	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	6142	0.40	ng	0.00
27) Chrysene-d12	21.40	240	7234	0.40	ng	0.00
34) Perylene-d12	23.90	264	6584	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.40	112	643	0.26	ng	0.01
5) Phenol-d6	7.02	99	559	0.16	ng	0.04
8) Nitrobenzene-d5	9.00	82	1442	0.36	ng	0.04
11) 2-Methylnaphthalene-d10	12.19	152	2866	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	488	0.35	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4763	0.41	ng	0.00
25) Fluoranthene-d10	19.24	212	39902	0.41	ng	0.00
29) Terphenyl-d14	19.84	244	7679	0.44	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.29	149	6390	0.133	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
Data File : BE098991.D
Acq On : 16 Mar 2019 6:47
Operator : JU/SJ
Sample : k1912-13RE
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
M3-T4-UV(POST-HYPO)RE

Quant Time: Mar 18 02:11:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

