

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
 Data File : BE098981.D
 Acq On : 15 Mar 2019 21:06
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 07:04:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.469	-17.2	135	0.00
3	n-Nitrosodimethylamine	0.400	0.303	24.3	86	0.04
4 S	2-Fluorophenol	0.400	0.318	20.5	89	0.02
5 S	Phenol-d6	0.400	0.332	17.0	96	0.01
6	bis(2-Chloroethyl)ether	0.400	0.306	23.5	92	0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.314	21.5	106	0.01
9	Naphthalene	0.400	0.392	2.0	122	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	122	0.00
12	2-Methylnaphthalene	0.400	0.372	7.0	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.400	0.269	32.8#	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	120	0.00
16	Acenaphthylene	0.400	0.371	7.3	124	0.00
17	Acenaphthene	0.400	0.370	7.5	119	-0.01
18	Fluorene	0.400	0.362	9.5	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.400	0.395	1.3	112	0.01
21	Hexachlorobenzene	0.400	0.390	2.5	114	0.00
22	Pentachlorophenol	0.400	0.373	6.8	69	0.00
23	Phenanthrene	0.400	0.375	6.3	107	0.00
24	Anthracene	0.400	0.324	19.0	95	0.00
25 SURR	Fluoranthene-d10	0.400	0.342	14.5	98	0.00
26	Fluoranthene	0.400	0.340	15.0	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
28	Pvrene	0.400	0.412	-3.0	97	0.00
29 S	Terphenyl-d14	0.400	0.386	3.5	91	0.00
30	Benzo(a)anthracene	0.400	0.360	10.0	86	0.00
31	Chrysene	0.400	0.382	4.5	92	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.321	19.8	82	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.348	13.0	83	0.00
34 I	Perylene-d12	0.400	0.400	0.0	96	0.00
35	Benzo(b)fluoranthene	0.400	0.375	6.3	93	0.00
36	Benzo(k)fluoranthene	0.400	0.372	7.0	94	0.00
37 C	Benzo(a)pyrene	0.400	0.397	0.8	98	0.00
38	Dibenzo(a,h)anthracene	0.400	0.314	21.5	76	0.00
39	Benzo(a,h,i)perylene	0.400	0.350	12.5	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
Data File : BE098981.D
Acq On : 15 Mar 2019 21:06
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 18 07:04:45 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				